

POLICE

HEADQUARTERS

J&K


SRINAGAR.

Subject: - Corrigendum in e-NIT No. 23 of 2016 dated 05.08.2016 floated by this Hqrs. for procurement of Hospital equipment.

Pursuant to queries raised during Pre-bid Conference held on 17.08.2016, specifications of item No.7 i.e, Computerized Perimeter tendered vide e-NIT No. 23 of 2016 dated 05.08.2016 read with Extension Notice issued under endstt. No. Prov-II/HE-04/16-17/52479-95 dated 07.09.2016 are hereby amended as under:-

S.No.	Name of the item with specifications	Tentative Qty.
01	Computerized Perimeter. <ul style="list-style-type: none">➤ Dimensions:<ul style="list-style-type: none">Height: Approx. 58 cmsWidth: Approx. 50 cms➤ Test specifications:<ul style="list-style-type: none">Maximum temporal range: 90°Stimulus duration: 200 msBackground illumination: 31.5 ASB➤ Threshold Testing:<ul style="list-style-type: none">24-2, 30-2, 10-2, 60-4, Macula, Nasal step➤ Threshold strategy:<ul style="list-style-type: none">SITA Standard:SITA FastFull Threshold➤ Fixation Control:<ul style="list-style-type: none">Heijl Krakua Blind Spot monitor.Video eye monitor.Gaze Tracking.Head Tracking.➤ Stimulus:<ul style="list-style-type: none">White on white.Red on Blue on White.Blue on Yellow➤ Testing Features:<ul style="list-style-type: none">Stimulus size: I-V Goldman's.Foveal Threshold testing.Automatic Pupil measurement.Liquid trails lens.➤ Software Features:<ul style="list-style-type: none">Glaucoma Hemifield Test (GHT).Visual Field Index (VFI).Guided Progressive Analysis (GPA).➤ Operator Interface:<ul style="list-style-type: none">Touch Screen CRT.Keyboard with mouse.Harddrive 500 GB Approx.USB.Printer.UPS.➤ Electrical Requirement:<ul style="list-style-type: none">100-240 V (50/60 HZ)	01 No.

Other terms and conditions of the NIT shall remain un-changed.


For Director General of Police,

J&K, Srinagar.

Dated:- 15 : 09.2016

No:Prov-II/HE-04/16-17/53420-40

Copy to the:-

1. Principal Secretary to Govt. Home Department J&K Srinagar for favour of information.
2. All ADsGP
3. All IsGP for information.
4. Director information J&K Srinagar for favour of information and with the request of arranging publication of corrigendum of e.tender in National Dailies viz, The Hindu, Decan Herald, Hindustan Times, Indian Express and in leading local news paper i.e Greater Kashmir in Srinagar and Excelsior in Jammu and Tender Journals etc. for wide publicity. The payment for publication in National Dailies shall be made by the PHQ against verified bills.
5. DIG CKR Srinagar/ DIG J-K Range Jammu
6. Director Industries Jammu/Srinagar for information and n/action.
7. Director Door Darshan Jammu/ Srinagar
8. Station Director Radio Kashmir Srinagar/ Jammu
,, for putting the corrigendum on AIR for three consecutive days.
9. Manager Govt. Press Srinagar/Jammu for publication of corrigendum of e-tender in Govt. gazette.
10. SSP J&K CID Cell, 11 Harish Chander Mathur Lane Kasturba Gandhi Marg New Delhi-110001 for information and n/action.
11. PRO to DGP J&K Srinagar alongwith 02 copies of the corrigendum of e-tender for publication purpose. He will personally liaise with Directorate Information for publication of the corrigendum immediately and make available concerned Newspaper in which corrigendum is published to AIG (Prov/Tpt) PHQ J&K for record and reference.
12. In-Charge PCS Gulshan Ground Jammu/ Zewan Sgr.
13. S.O e-procurement/~~In-Charge~~ IT Centre PHQ for placing corrigendum on the website of J&K Police and J&K Govt. Portal.



For Director General of Police,
J&K, Srinagar

9/15/9.