

**Subject :- Corrigendum/Extension in e-NIT No. 36 of 2015 dated 20.10.2015 floated by this Hqrs. for procurement of Hospital equipments.**

In view of some queries raised by intending firms, e-NIT No. 36 of 2015 dated 20.10.2015 read with Amendment Notice issued under endstt. No. Prov-II/HE-05/15-16/71330-50 dated 19.12.2015 specifications of item No.1 (Hysteroscope) are hereby further amended as under:-

S.No.	Name of the item with specifications	Tentative quantity
01	<p><b><u>HYSTEROSCOPE</u></b></p> <p><b><u>Hysteroscopy set:</u></b></p> <ol style="list-style-type: none"> <li>1. Examination sheath of suitable size with luer lock adapter.</li> <li>2. Operating sheath with instrument channel for operating hysteroscopy of suitable size.</li> <li>3. Autoclavable telescope 30deg. diameter 4mm, length 30cm.</li> <li>4. Polypectomy loop unipolar electrode suitable for polyductomy.</li> <li>5. Bipolar Dissection Electrode, semi rigid suitable for hysteroscopy.</li> <li>6. Monopolar High Frequency Cord with 4mm plug for HF unit, length 300cm</li> <li>7. Bipolar High Frequency cord with 4mm plug for HF Unit, length 300 cm</li> <li>8. Unipolar ball electrode suitable with above instruments.</li> <li>9. Flexible scissors single / double action jaws suitable with above instruments.</li> <li>10. Flexible Biopsy and grasping forceps double action jaws.</li> <li>11. Flexible Biopsy and grasping forceps double action jaws.</li> <li>12. Biopsy spoon forceps double action jaws.</li> </ol> <p><b><u>Irrigation Systems:</u></b></p> <ol style="list-style-type: none"> <li>1. Should be fully automatic.</li> <li>2. Should have irrigation capacity of 0 to 500 ml per minute</li> <li>3. Pressure range should be 10 – 200 mmHg</li> <li>4. Should have digital displays like preset pressure and preset flow</li> <li>5. Should be supplied with the accessories like silicon tubing compatible with the instruments supplied, power cord, etc.</li> <li>6. Should work with input 200 to 240Vac 50 Hz supply</li> </ol> <p><b><u>Digital endoscopic camera system:</u></b></p> <ol style="list-style-type: none"> <li>1. Should be a single chip camera technology.</li> <li>2. Should have one composite video outputs and one S-video output.</li> <li>3. Should have anti-moister filter for fiber scopes.</li> <li>4. Should have fully automatic exposure control.</li> <li>5. Should have automatic white balance with memory function.</li> <li>6. Should have horizontal resolution of more than 450 lines.</li> <li>7. Should be supplied with flat LCD TV of suitable size.</li> <li>8. Should work with input 200 to 240Vac 50 Hz supply</li> </ol>	02 Nos

**Light source and fiber optic light cable:**

1. Should be LED light source with minimum 150W light output.
2. Should have manual light intensity control.
3. Should have inbuilt cooling system.
4. Should be supplied with flexible fiber optic light cable with minimum diameter of 4.5 mm and minimum working length of 300cm.
5. Should work with input 200 to 240Vac 50 Hz supply.


**Others:**

- Should be supplied with suitable trolley.
  - Trolley should have at least 6 power sockets.
  - All equipment except irrigation system offered should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
- Comprehensive maintenance contract after expiry of warranty.

Accordingly, last date for submission/opening of online bids is hereby further extended as under:-

Last date for submission of online bids upto	04.01.2016 at 1800 hours
Date for opening of technical bids online	06.01.2016 at 1100 hours

The other terms and conditions of the NIT shall remain unchanged.

  
For Director General of Police,  
J&K, Jammu.  
Dated:- 23 : 09.2015

No:Prov-II/HE-03/15-16/72115-35  
Copy to the:-

1. Principal Secretary to Govt. Home Department J&K Jammu for favour of information.
2. All ADsGP
3. All IsGP for information.
4. Director General information J&K Jammu for favour of information and with the request of arranging publication of corrigendum of e.tender in National Dailies viz, The Hindu, Decan Herald, Hindustan Times, Indian Express and in leading local news paper i.e Greater Kashmir in Srinagar and Excelsior in Jammu and Tender Journals etc. for wide publicity. The payment for publication in National Dailies shall be made by the PHQ against verified bills.

5. DIG CKR Srinagar/ DIG J-K Range Jammu
6. Director Industries Jammu/Srinagar for information and n/action.
7. Director Door Darshan Jammu/ Srinagar
8. Station Director Radio Kashmir Srinagar/ Jammu  
,,, for putting the corrigendum on AIR for three consecutive days.
9. Manager Govt. Press Srinagar/Jammu for publication of  
corrigendum of e-tender in Govt. gazette.
10. SSP J&K CID Cell, 11 Harish Chander Mathur Lane Kasturba  
Ghandhi Marg New Delhi-110001 for information and n/action.
11. PRO to DGP J&K Jammu alongwith 02 copies of the corrigendum  
of e-tender for publication purpose. He will personally liaise with  
Directorate Information for publication of the corrigendum  
immediately and make available concerned Newspaper in which  
corrigendum is published to AIG (Prov/Tpt) PHQ J&K for record  
and reference.
12. In-Charge PCS Gulshan Ground Jammu/ Zewan Sgr.
13. S.O e-procurement/In-Charge ~~IT~~ Centre PHQ for placing  
corrigendum on the website of J&K Police and J&K Govt. Portal.

For Director General of Police,  
J&K, Jammu

23/12