

POLICE

HEADQUARTERS

J&K

JAMMU.


NIT NO: 10 of 2016

DATED: 4 /03/ 2016

For and on behalf of the Governor of Jammu and Kashmir State, e-tenders are hereby invited from original manufacturers or their authorized dealers having proper authorization of the manufacturer to quote against the tender for the supply of **Hospital Equipment** (as indicated in the detailed SBD & BOQs). The download of NIT shall start on **4 .3 .2016 at 1100 hours**. Tenders shall be submitted online on J&K State e-procurement Portal www.jktenders.gov.in before **16 .4 .2016 upto 1800 hours**. The samples (wherever asked) shall be submitted at Police Central Stores, Gulshan Ground, Gandhi Nagar, Jammu and Zewan, Srinagar before closing date of submission of e-tender. Samples received after the prescribed date/time shall not be accepted. **A Pre-bid conference shall be held on 16 .3 .2016 at 1500 hours in Conference Hall (4TH Floor) PHQ, J&K, Gulshan Ground, Jammu to clarify the queries of prospective bidders if any. All intending bidders should attend said conference to clarify matters related to this tender if any.**

The tenders (technical bids only) will be opened on-line on **18 .4 .2016 at 1100 hours at Police Headquarters J&K, Jammu**. In case of unforeseen circumstances, the date of opening will be next working day or else as notified separately.

Detailed tender documents (SBD) with terms and conditions are available on J&K State e-procurement Portal www.jktenders.gov.in


AIG (Provision/Transport)
For Director General of Police,
J&K, Jammu


No:Prov-II/HE- 17/2015-16/15430-50

Dated: 4 :03 : 2016

Copy to the:-

1. Principal Secretary to Govt. Home Department J&K Jammu for favour of information.
2. All ADsGP
3. All IsGP for information.
4. Director General Information J&K Jammu for arranging publication of Gist of the NIT in leading National Dailies like, The Hindu, Decan Herald, Hindustan Times, Indian Express and in leading local news paper i.e Greater Kashmir in Srinagar and Excelsior in Jammu and Tender Journals etc. for wide publicity. The payment for publication in National Dailies shall be made by the PHQ against verified bills.
5. DIG CKR Srinagar/ DIG J-K Range Jammu
6. Director Industries Jammu/Srinagar for information and n/action.
7. Director Door Darshan Jammu/ Srinagar

8. Station Director Radio Kashmir Srinagar/ Jammu
,, for putting the tender notice on air for three consecutive days.
9. Manager Govt. Press Srinagar/Jammu for publication of e-tender in Govt. gazette.
10. SSP J&K CID Cell, 11 Harish Chander Mathur Lane Kasturba Ghandhi Marg New Delhi-110001 alongwith 05 copies of the tender for information and distribution among intending firms.
11. Medical Supdt. PH, Jammu/ Srinagar
12. PRO to DGP J&K Jammu alongwith 06 copies of the gist of e-tender for publication purpose. He will personally liaise with Directorate Information for publication of the NIT/Gist immediately and make available concerned Newspaper in which NIT is published to AIG (Prov/Tpt) PHQ J&K for record and reference.
13. In-Charge PCS Gulshan Ground Jammu/ Zewan Sgr.
14. S.O. e-Procurement PHQ/ IT Centre PHQ for placing the e-tender on the website of J&K Police and J&K Govt. Portal.


AIG (Provision/Transport)
For Director General of Police,
J&K, Jammu
2/3

NIT NO. **10** of 2016
DATED. **4** /03/2016

SBD

1. For and on behalf of the Governor J&K State, online tenders(e-tenders) are invited from original manufacturers or their authorized dealers having specific authorization from their original manufacturer/s to quote against this tender for the supply of **Hospital Equipment** as detailed in Annexure 'A' to this NIT. The NIT consisting of qualifying information, eligibility criteria, specifications, bill of quantities (BOQ), set of terms and conditions of the contract and other details can be seen/downloaded from J&K State e-procurement portal www.jktenders.gov.in from **4.3.2016** at 1100 hours to **16.4.2016** upto 1800 hours.
Note;- Only those firms are allowed to respond who are either manufacturer/s or their authorized dealers. In case of authorised dealers, a specific authorization letter from the principal manufacturer with reference to this particular tender should be attached. Tenders of approved/authorized dealers without a specific authorization for this tender (as per format provided in appendix-1) from original manufacturers shall be rejected.
2.

Last date and time for receipt of samples (wherever asked)	16.4.2016 upto 1800 hours samples received after the prescribed date or time shall not be accepted. The samples shall be packed in a separate cover affixed with sealing wax and seal of the firm, quoting NIT No. and other details.
Address for submission of samples (wherever asked)	The samples shall be received only at Police Central Stores, Gandhinagar, Jammu or Police Central Stores, Zewan, Srinagar before closing date of submission of the tenders.
Time/date and place for online opening of technical bid only.	On 18.4.2016 at 1100 hours at J&K Police Hqrs. Gulshan Ground, Jammu.
Validity of tender offer	Upto 31.12.2016
3. Samples wherever asked shall be as under:
I. The tenderers shall ensure that sample submitted by them fully conforms to all the parameters of the NIT specifications.
II. Offers without samples after stipulated date and time for receipt of the tenders/samples shall not be accepted.
III. All samples should have a card affixed to it signed and stamped by the firm indicating (a) name of the firm (b) tender No. (c) date of opening of tender and (d) name of the item, make and model
IV. Sample shall be without any kind of marking/sign/symbol/blot/ scratch/strain. Sample not conforming to this clause shall be rejected.
V. Samples of rejected offers shall be collected by their authorized representatives on requisition made on a letter head of the firm, duly stamped, signed and attested by the tenderer. In case samples are not collected within the period of rate validity of this tender, the sample will be disposed off without any cost/liability to J&K Police.
4. All tender documents attached with this invitation to tender including the specifications are SACROSANCT for considering any offer as complete offer. It is important that scanned copy of each page of the TENDER DOCUMENT DULY completed, page numbered and signed is uploaded with your offer in a sequential manner and not haphazardly.
5. The tender fee has been fixed for SSI Units of J&K State only Rs. 100/- and all other vendors have to pay Rs 2000/- as tender fee in the shape of **crossed demand draft** favouring Chief Accounts Officer PHQ, J&K payable at Srinagar/Jammu. The tender fee is not refundable and the same will be deposited into Govt. Treasury under Head 0055-Police for the year 2015-16. The Scanned copy of the draft should be part of the online Bid. The original should be sent by post to the AIG (Provision/Transport)

c/o Director General of Police J&K Gulshan Ground Jammu Pin 180006 along with covering letter in an envelope should have a card affixed to it duly signed and stamped by the firm indicating (i) Name of the firm (ii) Tender No. (iii) Date of opening of the technical bids. The local SSI units of J&K State are required to pay the tender fee prescribed as per J&K Govt rules. However, the Govt./ Semi Govt. concerns of the J&K State claiming exemption shall append copies of the authenticated orders, existence/ registration certificates etc. with regards to the SSI Units issued by the Competent authority in support of their claim. Tenders without stipulated tender fee shall be rejected.

6. **Earnest Money Deposit (EMD)**

The intending tenderer/s will have to send the Earnest Money only in shape of CDR/FDR from any nationalized bank/ State Owned banks for an amount of Rs.5000/- for J&K State SSI Unit holders only and for other tenderers Rs. 50,000/- pledged to AIG (Prov/Tpt) PHQ J&K with the tender offer, which shall be valid for one year. The earnest money deposit in the case of unsuccessful tenderer shall be released after finalization of tender, where-as in the case of successful tenderers it will be adjusted towards the security money (required to be deposited) on application, for due performance of the agreed contract/ performance during the warranty period. No interest shall be payable by purchaser on the EMD deposited by the tenderer). The Scanned copy of the FDR/CDR should be part of the online Bid and the original be sent by post to the AIG (Provision/Transport) c/o Director General of Police J&K Gulshan Ground Jammu-180006 alongwith tender fee **before the opening of Technical Bid. J&K Govt. rules** with regard to furnishing of EMD in respect of local Registered SSI units shall be applicable in letter and spirit. The local SSI units are required to pay the EMD prescribed as per J&K Govt rules. For the purpose, SSI Unit shall have to furnish registration certificate INDICATING THEREIN THE DETAILS OF ITEMS AND INSTALLED CAPACITY OF THE UNIT PER MONTH AND FUNCTIONAL CERTIFICATE OF SSI Unit issued by the competent authority with their tender offer failing which the benefits of SSI Units will not be allowed. The Government/ Semi Govt. or similar other concerns of J&K State, claiming exemption from the payment of EMD shall have to append duly authenticated orders, certificate etc. from competent authority supporting their claims.

Tenders without EMD shall be rejected.

7.	i) Store required at	FREE DELIVERY AT CONSIGNEE LOCATIONS i.e, (Police Central Store Zewan, Srinagar and Police Central Store Gulshan Ground Gandhi Nagar Jammu). Tenderers are required to quote rates on free delivery at consignees locations basis only. Tenders received quoting rates on for destination stations/station of despatch basis shall be rejected.
	ii) Despatch instructions	Stores are required to be delivered at consignee locations as detailed in clause 7 (i).
	iii) Delivery required	Approximately within 45 days from the date of issue of supply order.
	iv) Inspection	The department is at liberty to conduct the inspections of stores during any stage of manufacturing. Tenderers should disclose the name and the full address (alongwith Telephone and fax No.) of the place where the stores will be manufactured and will be available for inspection.
	v) Inspection authority/place of survey	Board of officers, duly constituted by the J&K Police Hqrs. at consignees premises.
	vi) packing and marking	As per clause 29 of this tender
	vii) payment terms	Payment shall be released by the department against each consignment delivered, inspected, accepted (installed/commissioned wherever applicable) and properly brought on the stock in the concerned stock registers of the department. Remittance charges (bank commission) on payment made to the firms will be borne by the supplier/s. The preferred mode of payment will be through RTGS, for which the approved supplier shall have to furnish the Bank/RTGS code details. i. If approved supplier backs out or fails to supply the goods against supply order within stipulated delivery period, the department will be at liberty to procure the item from L-2 firm and differential

	<p>amount between L-1 & L-2 rates will be deducted from the security deposit furnished by the approved supplier and remaining security deposit will be forfeited besides invoking other remedies which shall include black-listing of the firm</p> <p>ii. If approved supplier supplies the goods partially against supply order within stipulated delivery period, J&K Police reserves the right to impose liquidity damage charges upto 5% of the un-supplied store upto first 15 days and thereafter penalty can be increased depending upon the nature of supplies and extent of delay. However, period of delivery may be extended at the absolute discretion of J&K Police provided delay is attributable to the reasons beyond the control of the supplier firm but not attributable to business transactions and also delay is communicated in writing justifying the cause of delay.</p>
8.	Rate quoted by the firms should be on firm price basis.
9.	Firm shall clearly indicate different taxes, duties, charges which they proposed to charge as per the format of Bill of Quantity (BOQ) available with the e-tender, <u>moreover online generation of financial comparative (BOQ Chart) is of indicative nature only and not conclusive. The detailed financial comparative (BOQ Chart) shall be worked out by the department and information can be uploaded on the website for general information of all the tenderers.</u>
10.	The approved supplier shall be deemed to have fully understood the conditions and specifications of the articles etc. to be supplied and in case of any confusion may seek clarification within 10 days from the date of uploading of tenders online and thereafter pre-bid conference can also be held by the J&K Police, if required.
11.	No conditional tender will be accepted/entertained. It will be summarily rejected. No other conditions than those mentioned (in the terms set-forth) will be entertained.
12.	Purchaser reserves the right to change the quantity of stores allotted to them.
13.	Purchaser reserves the right to get the manufacturing capacity of all the firms re-verified irrespective of the registration status or any other document/information.
14.	Purchaser reserves the right to cancel /reject or scrap the tender, with no consequential liabilities.
15.	<p><u>Two bid system;</u> All bidders are required to submit their offers in two covers online as under;- (A)<u>First cover (Technical Bid)</u> is to be uploaded online and scanned copies in PDF format should contain the following documents.</p> <ul style="list-style-type: none"> (i) The tenderer will clearly mention the name of the items for which he has quoted in the covering letter of the tender offer. (ii) Tender documents (viz. Technical Compliance, brochures/catalogues of the quoted items containing Make/Model of the item, OEM/Authorization Certificate, registration/existence certificates of the firms/SSI Units and performance etc) duly page numbered completed and signed BUT WITHOUT INDICATING THE RATES QUOTED. (iii) Scan copy of the Earnest money Deposit/CDR (iv) Scan copy of the tender fee deposit/DD. (v) Technical specifications of the items quoted with catalogues/brochures etc. and deposition of samples, (wherever asked). (vi) Test report of each sample (wherever asked) shall be complete in all respects. (vii) Tender with in-complete or without test report shall be rejected. (viii) Sample teste shall bear the seal of the testing laboratory prominently. (ix) The PHQ may at its discretion, before opening of financial bids, subject the test report/s to scrutiny and have the sample/s as received from the tenderers subject to fresh tests from independent source to check the veracity of the reports and in case it is found that sample does not conform to the BIS/IND & other specifications, such tender offer shall be rejected and J&K Police shall be at liberty to initiate legal proceedings against such

- firm.
- (x) Registration Certificate/Industrial License of original manufacturer, with due validity for the items quoted.
 - (xi) Certificate of being authorized dealer having proper and specific authorization (as per format provided in appendix-1) of the original manufacturer to quote against this tender.
 - (xii) Authorized dealer having authorization to quote against this tender, shall furnish his registration certificate with Central/ State Sale Tax Department for that particular item.
 - (xiii) Scan copy of the NIT, all pages duly signed and stamped by the tenderer, in token of having understood and accepted the specifications, terms and conditions of the NIT and quoted items duly ticked.
 - (xiv) Performance statement/turn over for last 02 years as appendix-III of this NIT.
 - (xv) Acceptance of the warranty/guarantee of the item.
 - (xvi) Acceptance regarding delivery within the stipulated delivery schedule.
 - (xvii) PAN No./VAT/TIN No. self attested photocopies thereof be appended.
 - (xviii) Check list for tenders as per appendix-III.
 - (xix) Any other relevant document which the firms wishes to submit

B. SECOND COVER (COMMERCIAL/PRICE BID) SHOULD CONTAIN THE FOLLOWING:

I.	Rate quoted by the firm should be on firm price basis as per bill of quantities (BOQs) of the e-tender to be downloaded from the website and Rates should be carefully quoted offline and then uploaded (Note); J&K Police does not issue road permits, concessional forms, state entry tax or octroi etc. exemption certificate, as such; Rate quoted shall be F.O.R consignee. The zero basic cost quoted for any item in BOQ will be treated as nil quoted.
II.	Rates quoted shall be final till the validity of the contract i.e, upto 31.12.2016 with a provision of further extension if finally approved by competent authority.

16. On due date of opening of tenders i.e **10 . 4 . 2016 at 1100 hours**, technical bids only shall be opened and the technical evaluation process will start. Those who qualify the technical bid will be informed online with the announcement of financial bid opening date.


17. Before opening of financial bids, samples of the technically qualified bidders may be further physically evaluated by J&K Police. The samples not conforming to the functional requirements of the purchaser/tender requirements shall be rejected. Decision of Police department regarding acceptance/rejection of the samples shall be final.

18. **Security deposit.** The approved tenderers will have to supply the articles as per approved rate, approved sample and specification till the validity of the contract. They will have to furnish the **security deposit @ 5%** of the total value of the items contracted for in the shape of CDR/FDR issued by a nationalized/scheduled bank, at the time of signing of **Deed of Agreement** to ensure satisfactory performance of the equipment supplied, for warranty period from the date of acceptance of product. This security deposit will be released in due course after the expiry of specified warranty period/agreed contract. The charges on stamps etc. (cost of documents in totality) shall be borne by the supplier while executing the agreement. (No interest shall be paid by the purchaser on the security deposit). J&K Govt rules for security deposit in respect of local SSI Units shall be applicable.

19. **Guarantee/warranty.**
Except as otherwise provided in the invitation to the tender, the contractor hereby declares that the goods, stores, articles sold/supplied to the purchaser under this contract shall be of the best quality and work man ship and new in all respects and shall be strictly in accordance with the specifications and particulars contained/mentioned in the contract. The contractor hereby guarantees that supplied goods/stores/articles would continue to conform to the specifications and quality for a period of 24 months from the date of acceptance. If during the period of 24 months supplied stores/good articles discovered not to be conforming to the specifications and quality or not giving satisfactory performance or have deteriorated, the decision of the purchaser in that behalf shall be final and binding on the contractor and purchaser shall be entitled to call upon the contractor to rectify/replace the goods/stores/articles or such portion thereof by the contractor and in such an event, the above mentioned warranty period shall apply to the goods/articles/stores rectified/replaced from the date of

	rectifications/replacement thereof. In case of failure of the contractor to rectify or replace with the goods etc. within specified time, the purchase shall be entitled to recover the cost with all expenses from the contractor for such defective stores by the available means.
20.	Liquidated Damages: The successful tenderers shall have to make the supplies within the stipulated period as would be mentioned in the supply order. J&K Police reserve the right to impose liquidity damages of 5% of cost of un-supplied stores which shall depend upon the nature of supplies and extent of delay. However, period of delivery may be extended, at absolute discretion of J&K Police provided the delay is attributable to reasons beyond the control of the supplier but not attributable to business transactions and provided delay is communicated in writing justifying the cause of delay.
21.	The supplies delivered by the successful tenderer shall be surveyed and verified by the J&K Police at Police Central Stores Jammu/Srinagar, as the case may be or after installation/commissioning of the same at the users end as the case may be.
22.	PHQ shall have the discretion to send any number of samples from the received supplies from successful bidder for testing to one or more NABL/BIS accredited labs to check its conformity with the approved sample/specifications quoted in the NIT/supply order. In case of variation, the supply will be rejected at the cost and consequences of supplier.
23.	As soon as the acceptance of the tender is communicated to the successful tenderers online, the contract shall be binding on him. The earnest money of the tenderer who backs out or withdraw his tender or fails to abide by it after acceptance thereof is communicated/posted to him shall be forfeited besides other penal remedies that may be available to the Govt. of the J&K under the law for the time being in force in the State of J&K;
24.	The successful tenderer shall have to enter into an agreement on prescribed proforma (immediately after the issue of supply order) with the Department for due performance of the contract which shall be invariably furnished before the supply is made.
25.	The tenderer is required to furnish an authenticated copy of Sales Tax/ VAT clearance certificates (alongwith the tender offer or at the time of payments) for the year 2014-15. The PHQ shall be at liberty to put to scrutiny, verification etc. the Sales tax /VAT clearance so furnished by the successful tenderer.
26.	The authorities reserve absolute powers to reject or accept any tender or part thereof without assigning any reason and without any legal obligations. Any clause incorporated in the tender (not conforming to the NIT in any manner) shall make such an offer liable for rejection. Therefore, all the terms and conditions of the tender Notice including technical specifications shall be carefully studied for the sake of complete and comprehensive tender. Failure to comply with any of the conditions stipulated in the tender or instructions or the offer with insufficient particulars/documents shall lead to out-right rejection of the tender offer.
27.	After the registration/placement of orders/execution of the agreement and formally depositing the security money, the tenderer backs out or fails to supply the goods, the Security Deposit will be forfeited besides any other penal provisions, the department may decide as per norms;
28.	The formal deed incorporating the terms of the contract will be executed by the successful tenderer immediately after receipt of supply order. Failure to execute such bond/deed shall not however prevent the contract from being enforced against the tenderer. Any loss sustained by the Govt. as a result of re-tendering the contract shall be recovered from the defaulter, besides other penal provisions as the department may decide;
29.	Delivery of Goods Goods shall be delivered (as per instructions in supply order) at Police Central Store, Gulshan Ground, Gandhi Nagar, Jammu and Police Central Store, Zewan, Srinagar as the case may be duly packed in perfect condition. The supplier if he so desires may insure the goods against loss or theft, destruction, damages by the natural calamity loss by exposure to weather or other wise viz war rebellion, riot etc. The insurance charges will have to be borne by the supplier and the department shall not pay such charges;
30.	If the successful tenderer fails to supply the goods of the prescribed specification or fails to deliver the goods within the stipulated period as specifically mentioned in the supply order, the department shall be at liberty to arrange supplies by re-tendering or other wise at the risk and cost of the tenderer. The department may give 15 days notice to the approved supplier in writing to make supplies of goods. In

	case of failure to supply the consignment on specified qualitative & quantitative terms, the department shall be lawfully entitled to forfeit the amount of the security money and take penal action to make good the loss sustained or excess cost incurred by the state in arranging of the supplies and other remedies that may be available to the department, under the law for the time being in force in the state. The imposable penalty can be 5% of the cost of goods depending upon nature of goods and cause of delay.
31.	Either original manufacturer or his authorised dealer can participate for a particular item or sample(s) only. Tender offers where both original manufacturer and his authorised dealer participate for the same items shall be rejected.
32.	If at any stage of tender process two or more bidders are found in collusion between/ among themselves with the intent to reduce competition, set price, share technical expertise, share business information/ decision, control/share production and or distribution of products and raise prices, all such tenders shall be out rightly rejected.
33.	If at any stage of tender process any bidder/s is/are found influencing any Government Servant associated with or handling the tender process and supplies thereof, bid of such tenderer/s shall be rejected besides invoking any other provision available under rules.
34.	All legal proceedings arising out of any dispute between the parties shall have to be settled in the Courts situated in Jammu and Srinagar and not elsewhere;
35.	It shall be incumbent upon the supplier to provide entire supply, in conformity with the approved sample, strictly as per the specifications provided in the NIT/supply order. Supplies found deficient in quality/quantity viz-a-viz supply order shall be rejected. Penalty/damages would be imposed and recovered from the bidder as per terms and conditions indicated in the supply order.
36.	The supplier shall not sublet the contract or any part thereof to any other agency.
37.	The decision of the accepting authority shall be final as to the quality of stores and shall be binding on the supplier. In case the articles supplied are not according to the approved sample & specifications they shall be rejected and at our absolute discretion, rejected stores may be allowed to be made good by way of replacement. The loss caused due to rejection of the supplies shall be entirely borne by the supplier. The rejected articles shall be lifted by the supplier within one week from the date of rejection. The department shall in no case be responsible for any loss or damage that may occur to the rejected stores while these are in the premises of the purchaser.
38.	The Department reserves the right to add or suitably modify the terms and conditions (without prior notification) as per exigencies and hence above shall not be construed as all inclusive.
39.	Bidders submitting their bids on-line shall only be accepted and manual bids shall be rejected and no document other than mentioned in the NIT shall be accepted manually.
40.	PHQ may call any bidder at any stage of the tender process for seeking any clarification.
41.	Queries/ representations on tender enquiry received during pre-bid conference or within 10 days from the date of publication of this tender enquiry (whichever is earlier) would be entertained. All queries/representations received thereafter would not be entertained.
42.	The bidder should furnish an undertaking APPENDIX-II of this NIT stating that he/ she is not connected in any manner with any other bidder in this tender with respect to manufacturing, distribution, supply, in sharing business information/documents price of the item etc. for the items quoted in this bid. If this is found to be incorrect, his/her tender would be rejected.
43.	The detailed NIT, SBD and BOQ is available on J&K State e-Procurement portal www.jktenders.gov.in . In case of any clarification, please contact on Phone no. 0191-2459029, 2459255


 (Anil Provision/Transport)
 For Director General of Police,
 J & K-Jammu.

APPENDIX-I
MANUFACTURER'S AUTHORIZATION LETTER/CERTIFICATE;

No,
The Director General of Police
J&K, Jammu

Subject:- _____

Sir,

Please refer to your e-tender enquiry (NIT) No- _____ of 2016 dated2016.

1. We _____ are proven and reputable Manufacturers of _____ (name and description of the goods offered in the tender) having factories at _____ hereby authorize M/s _____ (name and address of the agent) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred tender enquiry documents for the above said goods manufactured by us.
2. We also hereby extend our full warranty, CMC, support as per the conditions of contract, read with modifications, if any, in the special conditions of contract for the goods and services offered for supply by the above firm against this tender enquiry document.

Yours faithfully,

Signature with date, name and designation
For and on behalf of M/s _____
(Name and address of the Manufacturers)

Note:-

- a. This letter should be on the letter head of the manufacturing firm and should be signed by the person who is competent having the power of attorney.
- b. Scan copy of the original letter must be attached with tender documents.

SIGNATURE OF THE TENDERER.

APPENDIX-II

Reference eNIT No. of 2016 dated: : : 2016

The Director General of Police,
J&K, Jammu.

Sir,

1. I _____ hereby state that I am not connected in any manner with any other bidder in this tender with respect to manufacturing distribution, supply, in sharing business information/documents and price of the items etc. quoted in this bid.
2. I further state that if this is found to be correct, the Police department is at liberty to reject my tender.
3. I further certify that my concern has neither been blacklisted nor banned by any central or state organizations to participate in tendering process.

Yours faithfully,

Signature with date, name and designation
For and on behalf of M/s _____
(Name and address of the manufacturer)

Note:-

- a. This undertaking should be on the letter head of the manufacturer firm and should be signed by the authorized person.
- b. Scan copy of the original letter must be attached with tender documents.

SIGNATURE OF THE TENDERER.

APPENDIX-III

Performance statement for last two years

Name of the firm	
Description of stores	
Quantity on order	
Value	
Original D.P	
Quantity supplied within original D.P	
Final Ext D.P.	
Last supply position.	
Reasons for delay in supplies (if any)	

Signature of Tenderer

APPENDIX-IV

General instructions/CHECK LIST FOR TENDERERS;
Before submission of tender documents, the tenderers should check whether they have complied with the following requirements or otherwise;-

S.No	Requirement to be checked before submission of the tender	Complied (Please indicate YES or NO after complying with the requirement.
1.	The tenderers for their own ease shall fill in relevant details of NIT offline & upload the same after ensuring accuracy	
2.	Whether tender fee scan copy is enclosed, with the technical bid?	
3.	Whether Earnest Money Deposit (EMD) scan copy is enclosed with the technical bid?	
4.	If an SSI Unit of J&K State, the same is indicated in the tender and scan copy of the valid registration/functional certificate is enclosed or not?	
5.	Whether list of the quoted items with name, make, model and catalogue/brochures of the quoted items are uploaded with technical bid?	
6.	Whether scanned copies of complete tender documents are up-loaded in the technical bid duly signed having official seal on all pages or not. However, the quoted items may also be ticked ?	
7.	Whether tender documents are submitted in two bid system i.e. Technical and Financial Bids separately as per tender enquiry or not?	
8.	Wherever asked to deposit the samples, whether sample conforming to the specifications has been deposited with Central Store Jammu/Srinagar (as asked for in the specifications of item)?	
9.	Whether offer validity as required in tender enquiry is accepted and indicated in your tender document or not?	
10.	Whether terms of delivery and period as required in tender enquiry is accepted and indicated in your tender document or not?	
11.	Whether payment terms as required in the tender enquiry is accepted and indicated in your tender document or not?	
12.	Whether compliance statement in the format as required in tender enquiry is enclosed with the supporting catalogue/ brochures?	
13.	Whether performance statement for last 2 years as required in tender enquiry in the format is enclosed or not. If not, reason thereof is given?	
14.	Whether warranty period as required in the tender enquiry is accepted or not?	
15.	Whether details of AMC condition after warranty period (wherever applicable) is included in the price bid or not, if required in the tender enquiry?	
16.	(a) Status of tenderer as to whether manufacturer or authorized agent of the manufacturer is indicated in the tender document. If authorised agent/dealer, valid authority letter for the stores offered from the manufacturer is enclosed or not. (b) If an Indian agent of foreign principal, whether the firm is enlisted with DGS&D under compulsory enlistment scheme or not. Copy of valid DGS&D enlistment certificate as Indian Agent of Foreign Principal is enclosed or not. If not, reason thereof is given in tender document or not?	

17.	If called for in the T.E. whether acceptance of functional demo/field trial clause, (whenever asked) is specified in your tender document or not?	
18.	If called for T.E. whether free training about use of equipment after supply, is accepted or no?	
19.	Whether indicated in the tender document, if your firm's business dealings is banned by any Govt/Private agency or recommended for blacklisting or not?	
20.	The complete commercial quote should be in Indian Rupees only. Currency of price should be clearly mentioned (wherever applicable).	
21.	Tenderers are required to indicate Basic Rate, Custom Duty, Excise Duty, Sales Tax/VAT and any other leviable taxes/charges separately in the relevant columns of the BOQ.	
22.	Entry Tax/Octroi/Toll Tax and other Statutory Taxes will be applicable at actual supported by documents, if admissible	
23.	Tenderer will attach justification of price quoted which should include the recently executed contracts of the store in question with Govt/Semi Govt Organizations etc.	
24.	No column should be left blank. Wherever amount is not quoted, the column should be clearly marked as 'Nil'.	

List of sensitive and non sensitive documents.

S. No	Sensitive Documents part of Technical bid	Non Sensitive Documents
1.	Valid Registration Certificate/Industrial License of original manufacturer for the items quoted	Performance statement of two years
2.	Valid authorization letter of original manufacturer in case of approved dealer	Govt. supply orders for same if any
3.	Valid registration certificate of approved dealer with Central/State Sales Tax Department for the items quoted	Income tax clearance
4.	VAT clearance	
5.	SSI unit -registration certificate (local SSI units of J&K State only)	
6.	SSI unit – functional certificate (local SSI units of J&K State only)	
7.	Scanned copies of EMD and Tender Fee	
8.	Scanned copy of NIT duly Signed & sealed	

SIGNATURE OF THE TENDERER WITH SEAL

Annexure "A" to this Hqrs e.Tender No. 10 of 2016 dated . 4 . 03 . 2016

S.No.	Name of the item with specifications	Tentative quantity
01	<p><u>HYSTEROSCOPE</u> <u>Hysteroscopy set:</u></p> <ol style="list-style-type: none"> 1. Examination sheath of suitable size with luer lock adapter. 2. Operating sheath with instrument channel for operating hysteroscopy of suitable size. 3. Autoclavable telescope 30deg. diameter 4mm, length 30cm. 4. Polypectomy loop unipolar electrode suitable for polyductomy. 5. Bipolar Dissection Electrode, semi rigid suitable for hysteroscopy. 6. Monopolar High Frequency Cord with 4mm plug for HF unit, length 300cm 7. Bipolar High Frequency cord with 4mm plug for HF Unit, length 300 cm 8. Unipolar ball electrode suitable with above instruments. 9. Flexible scissors single / double action jaws suitable with above instruments. 10. Flexible Biopsy and grasping forceps double action jaws. 11. Flexible Biopsy and grasping forceps double action jaws. 12. Biopsy spoon forceps double action jaws. <p><u>Irrigation Systems:</u></p> <ol style="list-style-type: none"> 1. Should be fully automatic. 2. Should have irrigation capacity of 0 to 500 ml per minute 3. Pressure range should be 10 – 200 mmHg 4. Should have digital displays like preset pressure and preset flow 5. Should be supplied with the accessories like silicon tubing compatible with the instruments supplied, power cord, etc. 6. Should work with input 200 to 240Vac 50 Hz supply <p><u>Digital endoscopic camera system:</u></p> <ol style="list-style-type: none"> 1. Should be a single chip camera technology. 2. Should have one composite video outputs and one S-video output. 3. Should have anti-moister filter for fiber scopes. 4. Should have fully automatic exposure control. 5. Should have automatic white balance with memory function. 6. Should have horizontal resolution of more than 450 lines. 7. Should be supplied with flat LCD TV of suitable size. 8. Should work with input 200 to 240Vac 50 Hz supply <p><u>Light source and fiber optic light cable:</u></p> <ol style="list-style-type: none"> 1. Should be LED light source with minimum 150W light output. 2. Should have manual light intensity control. 3. Should have inbuilt cooling system. 4. Should be supplied with flexible fiber optic light cable with minimum diameter of 4.5 mm 	02 Nos

5. and minimum working length of 300cm.
Should work with input 200 to 240Vac 50 Hz supply.

Others:

- Should be supplied with suitable trolley.
 - Trolley should have at least 6 power sockets.
 - All equipment except irrigation system offered should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
- Comprehensive maintenance contract after expiry of warranty.

02

Short Wave Diathermy (CONTINUOUS/ PULSED)

02 nos

Applications:

- Effective in Inflammation of shoulder joint
- Elbow joint
- Cervical spondylosis
- Osteoarthritis
- Ligament sprains in knee joints
- Low back ache
- Heel pain
- Sinusitis etc.,

SPECIFICATIONS

1	Input Voltage:	100 – 240 V/ Max. 50/50 Hz
2	Output Frequency	27-12 MHZ
3	Wave length	11.00 meters
4	Output Power	1000 Watt Pulsed 500 Watt Continuous
5	Pulse Rate	5-800 pulses/ sec. 5, 20, 50, 100, 150, 200, 400, 800 Hz
6	Pulse width	20-400 Micro Seconds
7	RF Indicator	Available
8	Note	Built in heavy duty cooling fan for long life of value.
9	Weight	Up to 35 kgs.
10	Timer	Digital upto 30 min treatment time (termination of output & alarm at the end of treatment)
11	Distinctive ergonomic design with lockable castors	-
12.	Backlit display with clear operational information.	
13.	On screen simplicity for quick setup	
14.	10 programmable memory function atleast with optional treatment protocol	
15.	Accessories	Two 100 mm capacitive (Disc) Electrodes with

			durable electrode arm.	
03	<p align="center">Continuous Passive Motion (CPM)Unit (Lower/Shoulder)</p>			02 nos each
	Therapy Mode:	6 mobilization modes for continuous, intermittent and progressive passive movement.		
	ROM Setting:	Programmable from 0 to 110°		
	Pause Time:	Programmable extension, flexion, from 0 to 90 sec.		
	Treatment time:	Programmable from 1 to 99 min.		
	Inching Switch	Starting angle can be set according to need of patient.		
	Memory/ Program	30 treatment protocols can be programmed and stored in memory		
	Display	Digital display of ROM setting, Pause time/ treatment time/ graphic display of mobilization mode.		
	Patient safety	Patient stop/ interruption treatment with handheld switch if needed		
	Trolley	Provides transportation and storage		
	Power supply	200 – 240 V A/C at 50-60 Hz.		
	<u>SHOULDER</u>			
	Therapy Mode:	6 mobilization modes for continuous, intermittent and progressive passive movement.		
	ROM Setting:	Programmable from 0 to 180° & 90 - + 90° for all shoulder movements.		
	Pause Time:	Programmable extension, flexion, from 0 to 99 sec.		
	Treatment time:	Programmable from 1 to 99 min.		
	Inching Switch	Starting angle can be set according to need of patient.		
	Memory/ Program	80 preset treatment protocols stored in memory		
	Open Programme	20 Programme space free to store user profiles/ patterns		
	Display	Digital display of ROM setting, Pause time/ treatment time/ graphic display of mobilization mode.		
	Height Alignment	Electrical by push button switch		
	Arm Setting	Left or right arm can be set and exercised by little adjustment.		
	Patient safety	Patient stop/ interruption treatment with handheld switch if needed		
	Power supply	200 – 240 V A/C at 50-60 Hz.		
04	<p align="center"><u>ELECTROLYTE ANALYSER</u></p>			02 nos
	<ol style="list-style-type: none"> 1. Ion Selective Electrode (ISE) for Sod., Pot., and Chloride 2. Maintenance free electrodes. 3. 45 to 60 samples per hour. 4. User programmable sample volume from 90 – 200 ul. 5. Low reagent consumption. 6. Should have alpha numeric display. 7. Automatic probe cleaning. 			

	<ol style="list-style-type: none"> 8. Simple and reliable design of fluid tubing system and electrodes. 9. Must be provided with electrode activator and refilling solution. 10. Internal build-in thermal printer. 11. Large 5.6 inch touch screen LCD with back light. 12. Two Calibration modes (Automatic/ Manual) 13. Should have facility to store upto 5000 test results. 14. Automatic warning reagent replacement and inadequate sample. 15. Able to provide 13 digit sample ID. 16. In built QC programme. 17. Should be supplied with On line UPS with minimum 30 minutes backup. 18. Should be provided with calibration certificate issued by the manufacturer. 19. Should be CE/FDA Approved. 20. Comprehensive maintenance contract after expiry of warranty. 	
05	<p style="text-align: center;"><u>BINOCULAR MICROSCOPE</u></p> <ol style="list-style-type: none"> 1. Optic System: UIS 2 (Universal Infinity System) 2. Illumination System: Built-in Transmitted illumination system, LED 3. Focusing: Stage height movement (coarse movement stroke 20 mm) Fine focus graduation 2.5 m 4. Revolving Nose piece: Fixed quadruple nosepiece. 5. Stage: Wire movement mechanical fixed stage 120X132 mm Travelling range: 76 mm (X) x 30 mm (Y) 6. Observation Tube: 30° inclined binocular tube. 7. Condenser: Abbe Type with aperture iris diaphragm N.A. 125 8. Objectives: Plan Achromatic objective (Anti-fungal) <ol style="list-style-type: none"> 4X N.A.: 0.10 W.D.: 18.5 mm 10X N.A.: 0.25 W.D.: 10.6 mm 40X N.A.: 0.65 W.D.: 0.06 mm 100XOil: N.A.: 1.25 W.D.: 0.13 mm 9. Eye Piece (10X): Field Number (F.N.), 20 (Anti-Fungus) 10. Optional Accessories: Mirror unit, 15X eye piece (F.N. 12, anti fungus) Micrometer, Filter holder, dark field Stop, micro image projection system, Phase-contrast Attachment/ 10X, 40X, Simple Polarizing attachment. 11. Weight: Approx. 6 Kgs. 12. Rated Voltage/ Electric Current: AC 100-240 V, 50/60 Hz 0.4 A 13. Power consumption: 17 W 14. Comprehensive maintenance contract after expiry of warranty. 	01 No.
06	<p style="text-align: center;"><u>SUCTION APPARATUS</u></p> <ol style="list-style-type: none"> 1. Should be approved from Bureau of Indian Standards. 2. Anti-corrosive, Powder Coated Chassis 3. 200 Watt Heavy Thermal Protector Electric Motor 4. 2 X 2.0 Ltrs. Polycarbonate jars with overflow safety 5. Noise level of suction apparatus from is 50 dB +/- 03 dB 6. Electrical Requirement - 220~230V, 50Hz, 1Phase 7. Ideal for MTP / medical / surgical procedures 8. Oil Free Rocker Piston Pump with -750 +/- 10 mmHg 9. Free air displacement >100 LPM 10. Non collapsible Silicon tubing 11. Bacterial filter fitted on top 12. Heavy duty HN-65 PU Castors 13. 100 mm Vacuum Gauge 14. Minimum Noise production. 	04


07	<p style="text-align: center;"><u>Foldable Stretcher</u></p> <p>Folding stretcher/ foldable stretcher. Having High quality, great service. 3. Should be easy to operate, very convenient stain.</p> <p>Features:</p> <ul style="list-style-type: none"> -Folding stretcher made of best quality fabric sturdy with folding frame, -Should have heavy metal pipe on the both sides and cross strip provided on the both side of the stretcher to open or to close the stretcher. -Rubber handle to carry the unit. -Light weight easily foldable. -Fabric quality tatron material with stretcher size Approx. 22"x 84" or more with pole size of Approx. 7 ft each side four heavy metal strap. -U shape fit provided to hold the stretcher above from the earth. 	04
08	<p style="text-align: center;">Micro Wave Therapy Unit</p> <p>Tech. Specifications:</p> <ul style="list-style-type: none"> • 250 Watt in constant mode. • 1600 Watt in Pulse Mode. • Overheating alarm. • Double protection system to guaranty extended magnetron life integrated. • Serial controller 16 bit or more microprocessor. • Approx. 100 memorized protocols. <p>Space for memorizing 100 protocols.</p>	01
09	<p style="text-align: center;"><u>HAEMATOLOGY ANALYSER</u></p> <ol style="list-style-type: none"> 1. Should be fully automated 5 part differential haematology analyser based on flow cytometer, light scattering and peroxidase staining technology. 2. Should have five discrete analysis modes CBC, CBC+DIFF, CBC+Retic, DIFF & Retic only. 3. Should give 33 or more parameters viz. WBC, RBC, HGB, HCT, MCV, MCH, MCHC, CHCM, RDW, HDW, CH, CHDW, Plt, MPV, PDW, PC, T% RETIC, # RETIC MCV, CHCM, RDW, HDW, CH, CHDW, A Solute & % value for NEUT, LYMPH, MONO, EOS, BASO, LUC, morphology results (User definable) like WBC left shift typical Lymph, Immature, Granulocytes, RBC, NRBC, ANISO, CICRO, MACRO, RBC Ghosts. 4. Should have capacity to perform veterinary application with minimum of 21 pre programmed species & 30 programmable slots. 5. Should have an Auto sampler with capacity of 150 sample tubes at a give time. A single sample rack should be able to cater different tube size . 6. Should have a high through put 120 samples per hour or more in CBC and CBC/ Diff mode & 74 samples per hour or more in Retics mode. 7. Should have multichannel analysis for better resolution and reproducibility like Dual differential count for WBC, Platelets/ RBC should have light scatter, HGB should have photometric and direct cellular, Retics should have on board light scatter (fluorescent dye) for Reticulocytes. 8. Should have on board reagent facility with maximum on board and automatic reagent inventory management. 9. Should have extensive linearity as WBC-0.02-400x10³, TBC0.7x10⁶, Plts.0-3500x10³ HGB 0-22.5 g/dl TETIC 0.2-24.5%. 10. Should have FDA approved capability of running CSF fluids, carryover of <or= to 1% for all parameters, sample volume required in all modes not exceed 175 ul dead volume required <300 ul. 11. Should be capable of reporting unique additional parameters such as CHr-reticulocytes HGB measurement to aid in anemia treatment. 12. Should have extensive QC features 3D Bar & SDI Graph LJ plot available table format delta checks available for cumulative review option for online QC also available patient moving average QC file management ILQC Programme remote QC. 13. Should have comprehensive data management such as user friendly windows based software, network integration possible with alb information system database, storage capacity of 10,000 records including graphics. 14. USFDA Approved. 	02

10	<p><u>RETINO SCOPE (HEINE)</u></p> <ul style="list-style-type: none"> ➤ Should have multicoated optics with parastop ➤ Ergonomic shape ➤ XHL Xenon Halogen Technology ➤ Metal Controls for durability dustproof ➤ Integrated polarization filter with orange filter ➤ Detachable brow rest ➤ Holder for fixation cards 	01
11	<p><u>SIX MIRROR VOLK GONIO LENS</u></p> <ul style="list-style-type: none"> ➤ Should have 6 mirrors to eliminate gaps ➤ Should provide 360 degree panoramic view of the fundus ➤ Tapered profile for patient comfort ➤ Taller, narrower body ➤ No flange/ fluid design 	01
12	<p><u>COMPUTERIZED PERIMETER</u></p> <ul style="list-style-type: none"> ➤ Size 24" H & 19Dx23"W ➤ Weight- 80 lbs ➤ Amperage- 4A ➤ Bowl illuminator – 31.5 ASB ➤ Frequency – 50-60Hz ➤ Humidity- 20-85% ➤ Maximum horizontal range – 86 degree ➤ Maximum stimulus intensity – 10,000 ASB ➤ Max wet bulb -78F ➤ Operating temperature 41-95 F ➤ Power supply voltage – 120/230 V AC ➤ Stimulus duration – 200 msec <p>Stimulus wave length- 30 cm</p>	01
13	<p><u>PHACO TUBING WITH TIP</u></p> <ul style="list-style-type: none"> ➤ Made up of silicon and not by PVC related polymer <p>It should be autoclavable and reusable</p>	01
14	<p><u>BLOOD CLOTTING PATCH</u></p> <ul style="list-style-type: none"> ➤ Should be sterile, non-absorbable Chitosan based hemostatic dressing pads ➤ Should be able to control severe bleeding in 2-3 minutes ➤ Should be usable both externally as well as internally ➤ Should provide an active mechanical barrier at the wound site ➤ Should not require premixing or preparation prior to use ➤ Should be completely removable from wound without dislodging the already formed clot ➤ Should be X-ray compatible <p>Should be CE mark</p>	200
15	<p><u>COLOUR DOPPLER</u></p> <p><u>TECHNICAL SPECIFICATION</u></p> <p>LATEST GENERATION HIGH END ULTRASOUND EQUIPMENT WITH COLOUR DOPPLER & 3D-4D FACILITY</p> <p>The system should meet the following minimal requirements and specifications:-</p> <p><u>01. PRESET EXAMINATIONS</u></p>	01

- ✔ Obstetrics & Gynaecology
- ✔ Abdomen
- ✔ Vascular
- ✔ Small parts-Breast-Scrotum, Thyroid etc.
- ✔ Urology
- ✔ Musculoskeletal
- ✔ Intracavitary -Trans vaginal
- ✔ 3D & 4D
- ✔ Elastography with latest Shearwave technology on convex and linear probes
- ✔ The equipment should be contrast agent imaging upgradable
- ✔ The unit must be upgradable to Fusion Imaging in future
- ✔ Touch screen control panel
- ✔ Gel warmer
- 02 MODES**
- ✔ 2-D Mode
- ✔ M Mode
- ✔ Pulsed Wave Mode
- ✔ Colour Flow Imaging
- ✔ Continuous Wave Mode
- ✔ Colour Power Angio/ Power Doppler Imaging
- ✔ Tissue Harmonic Imaging
- ✔ Duplex & Triplex Mode simultaneity (all three modes 2D, Colour & Doppler Modes simultaneously)
- ✔ Anatomical M mode
- ✔ Tissue Doppler imaging
- 03 TRANSDUCERS: Broad bandwidth for extreme high resolution 2D imaging**
- ✔ Convex Array Transducer-Frequency 2-5 MHz (+/- 1MHz) with 30 cm (or more) scan depth
- ✔ 3D-4D TVS probe
- ✔ Linear Array Transducer-Frequency 4-10 MHz(+/-1MHz)
- ✔ Volume convex transducer – 2-6 MHz (+/-1MHz)
- 04 BIOPSY ATTACHMENTS/NEEDLE GUIDE & BIOPSY/NEEDLE GUIDE SOFTWARE WITH ALL PROBES (EXCEPT CARDIAC & VOLUME PROBE)**
- ✔ **PORTS:** Atleast 3 active ports+ 1 parking prot/slot for transducers
- 05 APPLICATIONS AND FEATURES:**
- ✔ Fast boot-up time (for start-up to scanning)
- ✔ Sturdy, resistant to breakage & damage on fall/ hit against the wall/hard surface
- ✔ New generation ergonomically designs to curve minimum injury/ strain to the operator
- ✔ Dual &/ Quad split display
- ✔ Duplex for simultaneous B mode & spectral Doppler
- ✔ Facility of full screen imaging
- ✔ Full screen display of all 3D views including individual X,Y,Z MPR views real time acquisition and display of two image planes simultaneously with colour
- ✔ Facility for zoom (real-time and frozen image)
- ✔ Cineloop review facility in individual and mixed modes with memory upto minimum of 2000 images and 60 seconds of M Mode date.
- ✔ One touch image optimization & automatic real-time Doppler tracing

- 3D/4D features like MPR, Curved Rendering, Image manipulations of all modes including pre & post processing with 4D cine loop review
- **Elastography with quantification tools** including strain measurement, shell measurement, Histogram measurement (with Linear & convex transducers) with shearwave technology.
- Controls for time Gain Compensation (TGC-slide pot control), variable power, dynamic range & overall gain control so that following parameters can be adjusted on stored/ live image-Gain, TGC, dynamic range, Gray map, Colour map, colour gain, Doppler gain & Doppler angle correction
- Minimum 55000 **digital processing channels** for high resolution imaging with high acquisition frame rate (900 fps or more)
- **170 dB or more full time input dynamic range**
- **Gray Scale- Minimum 256 or more**
- **Pan zoom facility on live and freeze image**
- **Tissue Harmonic Imaging capability on phased, Linear, 3D and curved array transducers. Tissue Harmonic Imaging should be available in colour flow imaging, M-Mode and 3D rendering modes**
- **Powerful multi beam parallel imaging**
- **Compound Imaging**
- **Extended field of view imaging**
- **Trapezoid imaging on linear probe**
- **Panoramic imaging**
- **Pan Zoom feature for real time and Frozen images**
- **Image display-Dual & Quad**
- **Independent steering of B mode and Colour beam on linear probe**
- **Speckle reduction technology**
- **High resolution, non-interlaced, flicker free, antiglare, LCD/LED medical grade flat panel display of 19" or more, with maximum viewing angle & provision to adjust height**
- **Alpha numeric key board with track ball – fully articulating control Panel including height, swivel & alide adjustments and full alphanumeric keyboard with illuminated keys**
- **Complete package for measurement & calculation for all preset examinations mentioned at Sr. No. 1, with digital calipers for distance, area, volume, circumference & Doppler wave from parameters measurements etc**
- **PW Doppler analysis in Auto & Manual Mode, Variable sample volume, selectable wall filter, Angle correction**
- 06 **IMAGE MANAGEMENT FACILITY:**
- **Inbuilt hard disk with memory of minimum 500 GB**
- **Direct storage of Images & Cine loops in Hard disk**
- **Thumbnail review to view & edit images, loops & reports**
- **Display annotations**
- **Patient id display**
- **Provision for reverse & invert facility**
- **Image archival & Transfer**
- **Inbuilt CDRW, USB Drive with the facility to transfer image**
- **USB port (minimum one)**
- **Lan Port/ Ethernet port (or wireless LAN Support)**
- **DICOM, 3 Compatible with Life time DICOM Licence**
- **Upgradability for PACS**

	<ul style="list-style-type: none"> ➤ Direct Connectivity to Laser/ Thermal Printer for printing images & reports 	
07	<u>BATTERY BACKUP:</u>	
	➤ Battery back-up of atleast 30 min with on-line UPS	
08	<u>STANDARD ACCESSORIES</u>	
	➤ B/W thermal printer or latest model (with CE or FDA mark)	
	➤ Latest model colour laser printer	
	➤ Free software upgrade (s) during the period of warranty & CMC	
	➤ Probe protection covers/ probe connector cover for all probes	
	➤ On-line UPS of appropriate rating with 60 mins back up	
	➤ Ability to perform well in temp range from 5-40 degree Celcius & Relative humidity of 30-85%	
09	<u>OTHERS</u>	
10	Comprehensive hands- on training in a reputed institute for technical person (Radiologist/ Sonologist) and support services till familiarity with the system.	
11	US FDA & CE European approved. Requisite certificate to be attached	
Warranty for 2 years + 5 year CMC		


 AIG/ (Provision/Transport)
 For Director General of Police,
 J & K, Jammu