


NIT NO: 28 of 2016  
DATED: 12/09/2016

For and on behalf of the Governor of Jammu and Kashmir State, e-tenders are hereby invited from original manufacturers or their authorized dealers having proper authorization of the manufacturer to quote against the tender for the supply of **Hospital Equipment** (as indicated in the detailed SBD & BOQs). The download of NIT shall start on 12.09.2016 at 1100 hours. Tenders shall be submitted online on J&K State e-procurement Portal [www.jktenders.gov.in](http://www.jktenders.gov.in) before 12.10.2016 upto 1800 hours. The samples (wherever asked) shall be submitted at Police Central Store, Zewan, Srinagar before closing date of submission of e-tender. Samples received after the prescribed date/time shall not be accepted. **A Pre-bid conference shall be held on 09.09.2016 at 1500 hours in Conference Hall (3<sup>rd</sup> Floor) PHQ, J&K, Srinagar to clarify the queries of prospective bidders, if any. All intending bidders should attend said conference to clarify matters related to this tender, if any.**

The tenders (technical bids only) will be opened on-line on 14.10.2016 at 1100 hours at Police Headquarters J&K, Srinagar. In case of unforeseen circumstances, the date of opening will be next working day or else as notified separately.

Detailed tender documents (SBD) with terms and conditions are available on J&K State e-procurement Portal [www.jktenders.gov.in](http://www.jktenders.gov.in)


  
AIG (Provision/Transport)  
For Director General of Police,  
J&K, Srinagar  
9/9

No:Prov-II/HE-01/2014-15/ 5324-61  
Dated: 12:09 : 2016

Copy to the:-

1. Principal Secretary to Govt. Home Department J&K Srinagar for favour of information.
2. All IsGP for information.
3. Director Information J&K Srinagar for arranging publication of Gist of the NIT in leading National Dailies like The Hindu; Decan Herald, Hindustan Times, Indian Express etc. and in leading local news paper i.e Greater Kashmir in Srinagar and Excelsior in Jammu and Tender Journals etc. for wide publicity. The payment for publication in National Dailies shall be made by the PHQ against verified bills.
4. DIG CKR Srinagar/ DIG J-K Range Jammu
5. Director Industries Jammu/Srinagar for information and n/action.
6. Director Door Darshan Jammu/ Srinagar

7. Station Director Radio Kashmir Srinagar/ Jammu  
,, for putting the tender notice on air for three consecutive days.
8. Manager Govt. Press Srinagar/Jammu for publication of e-tender in Govt. gazette.
9. SSP J&K CID Cell, 11 Harish Chander Mathur Lane Kasturba Ghandhi Marg New Delhi-110001 for information.
10. Medical Supdt. PH, Jammu/ Srinagar
11. PRO to DGP J&K Srinagar alongwith 06 copies of the gist of e-tender for publication purpose. He will personally liaise with Directorate Information for publication of the NIT/Gist immediately and make available concerned Newspaper in which NIT is published to AIG (Prov/Tpt) PHQ J&K for record and reference.
12. In-Charge PCS Gulshan Ground Jammu/ Zewan Sgr.
13. S.O. e-Procurement PHQ/ IT Centre PHQ for placing the e-tender on the website of J&K Police and J&K Govt. Portal.

  
AIG (Provision/Transport)  
For Director General of Police,  
J&K, Srinagar  
21/9

NIT NO. <sup>28</sup> of 2016  
DATED. <sup>19</sup>. 09. 2016

**SBD**

1. For and on behalf of the Governor J&K State, online tenders(e-tenders) are invited from original manufacturers or their authorized dealers having specific authorization from their original manufacturer/s to quote against this tender for the supply of **Hospital Equipment** as detailed in Annexure 'A' to this NIT. The NIT consisting of qualifying information, eligibility criteria, specifications, bill of quantities (BOQ), set of terms and conditions of the contract and other details can be seen/downloaded from J&K State e-procurement portal [www.jktenders.gov.in](http://www.jktenders.gov.in) from 12.09.2016 at 1100 hours to 12/10.2016 upto 1800 hours.
- Note;- Only those firms are allowed to respond who are either manufacturer/s or their authorized dealers. In case of authorised dealers, a specific authorization letter from the principal manufacturer with reference to this particular tender should be attached. Tenders of approved/authorized dealers without a specific authorization for this tender (as per format provided in appendix-1) from original manufacturers shall be rejected.**
2. 

Time/date and place for online opening of technical bid only.	On 19.10.2016 at 1100 hours at J&K Police Hqrs. Peerbagh, Srinagar.
Validity of tender offer	Upto 30.12.2017
3. Samples wherever asked shall be as under;
- I. The tenderers shall ensure that sample submitted by them fully conforms to all the parameters of the NIT specifications.
  - II. Offers without samples after stipulated date and time for receipt of the tenders/samples shall not be accepted.
  - III. All samples should have a card affixed to it signed and stamped by the firm indicating (a) name of the firm (b) tender No. (c) date of opening of tender and (d) name of the item, make and model
  - IV. Sample shall be without any kind of marking/sign/symbol/blot/ scratch/strain. Sample not conforming to this clause shall be rejected.
  - V. Samples of rejected offers shall be collected by their authorized representatives on requisition made on a letter head of the firm, duly stamped, signed and attested by the tenderer. In case samples are not collected within the period of rate validity of this tender, the sample will be disposed off without any cost/liability to J&K Police.
4. All tender documents attached with this invitation to tender including the specifications are SACROSANCT for considering any offer as complete offer. It is important that scanned copy of each page of the TENDER DOCUMENT DULY completed, page numbered and signed is uploaded with your offer in a sequential manner and not haphazardly.
5. The tender fee has been fixed for SSI Units of J&K State only Rs. 100/- and all other vendors have to pay Rs 2000/- as tender fee in the shape of **crossed demand draft** favouring Chief Accounts Officer PHQ, J&K payable at Srinagar/Jammu. The tender fee is not refundable and the same will be deposited into Govt. Treasury under Head 0055-Police for the year 2016-17. The Scanned copy of the draft should be part of the online Bid. The original should be sent by post to the AIG (Provision/Transport) c/o Director General of Police J&K, Perbagh, Srinagar Pin 190014 along with covering letter in an envelope should have a card affixed to it duly signed and stamped by the firm indicating (i) Name of the firm (ii) Tender No. (iii) Date of opening of the technical bids. The local SSI units of J&K State are required to pay the tender fee prescribed as per J&K Govt rules. However, the Govt./ Semi Govt. concerns of the J&K State claiming exemption shall append copies of the authenticated orders, existence/ registration certificates etc. with regards to the SSI Units issued by the Competent authority in support of their claim. Tenders without stipulated tender fee shall be rejected.
6. **Earnest Money Deposit (EMD)**  
The intending tenderer/s will have to send the Earnest Money only in shape of CDR/FDR from any nationalized bank/ State Owned banks for an amount of Rs.5000/- for J&K State SSI Unit holders only

and for other tenderers Rs. 50,000/- pledged to AIG (Prov/Tpt) PHQ J&K with the tender offer, which shall be valid for one year. The earnest money deposit in the case of unsuccessful tenderer shall be released after finalization of tender, where-as in the case of successful tenderers it will be adjusted towards the security money (required to be deposited) on application, for due performance of the agreed contract/ performance during the warranty period. No interest shall be payable by purchaser on the EMD deposited by the tenderer). The Scanned copy of the FDR/CDR should be part of the online Bid and the original be sent by post to the AIG (Provision/Transport) c/o Director General of Police J&K Peerbagh, Srinagar-190014 alongwith tender fee **before the opening of Technical Bid. J&K Govt. rules** with regard to furnishing of EMD in respect of local Registered SSI units shall be applicable in letter and spirit. The local SSI units are required to pay the EMD prescribed as per J&K Govt rules. For the purpose, SSI Unit shall have to furnish registration certificate INDICATING THEREIN THE DETAILS OF ITEMS AND INSTALLED CAPACITY OF THE UNIT PER MONTH AND FUNCTIONAL CERTIFICATE OF SSI Unit issued by the competent authority with their tender offer failing which the benefits of SSI Units will not be allowed. The Government/ Semi Govt. or similar other concerns of J&K State, claiming exemption from the payment of EMD shall have to append duly authenticated orders, certificate etc. from competent authority supporting their claims.

**Tenders without EMD shall be rejected.**

7.	i) Store required at	FREE DELIVERY AT CONSIGNEE LOCATIONS i.e. (Police Central Store Zewan, Srinagar and Police Central Store Gulshan Ground Gandhi Nagar Jammu). Tenderers are required to quote rates on free delivery at consignees locations basis only. Tenders received quoting rates on for destination stations/station of despatch basis shall be rejected.
	ii) Despatch instructions	Stores are required to be delivered at consignee locations as detailed in clause 7 (i).
	iii) Delivery required	Approximately within 45 days from the date of issue of supply order.
	iv) Inspection	The department is at liberty to conduct the inspections of stores during any stage of manufacturing. Tenderers should disclose the name and the full address (alongwith Telephone and fax No.) of the place where the stores will be manufactured and will be available for inspection.
	v) Inspection authority/place of survey	Board of officers, duly constituted by the J&K Police Hqrs. at consignees premises.
	vi) packing and marking	As per clause 29 of this tender
	vii) payment terms	<p>Payment shall be released by the department against each consignment delivered, inspected, accepted (installed/commissioned wherever applicable) and properly brought on the stock in the concerned stock registers of the department. Remittance charges (bank commission) on payment made to the firms will be borne by the supplier/s. The preferred mode of payment will be through RTGS, for which the approved supplier shall have to furnish the Bank/RTGS code details.</p> <p>i. <b>If approved supplier backs out or fails to supply the goods against supply order within stipulated delivery period, the department will be at liberty to procure the item from L-2 firm and differential amount between L-1 &amp; L-2 rates will be deducted from the security deposit furnished by the approved supplier and remaining security deposit will be forfeited besides invoking other remedies which shall include black-listing of the firm</b></p> <p>ii. <b>If approved supplier supplies the goods partially against supply order within stipulated delivery period, J&amp;K Police reserves the right to impose liquidity damage charges upto 5% of the un-supplied store upto first 15 days and thereafter penalty can be increased depending upon the nature of supplies and extent of delay. However, period of delivery may be extended at the absolute</b></p>

discretion of J&K Police provided delay is attributable to the reasons beyond the control of the supplier firm but not attributable to business transactions and also delay is communicated in writing justifying the cause of delay.

8. Rate quoted by the firms should be on firm price basis.
9. Firm shall clearly indicate different taxes, duties, charges which they proposed to charge as per the format of Bill of Quantity (BOQ) available with the e-tender. moreover online generation of financial comparative (BOQ Chart) is of indicative nature only and not conclusive. The detailed financial comparative (BOQ Chart) shall be worked out by the department and information can be uploaded on the website for general information of all the tenderers.
10. The approved supplier shall be deemed to have fully understood the conditions and specifications of the articles etc. to be supplied and in case of any confusion may seek clarification within 10 days from the date of uploading of tenders online and thereafter pre-bid conference can also be held by the J&K Police, if required.
11. No conditional tender will be accepted/entertained. It will be summarily rejected. No other conditions than those mentioned (in the terms set-forth) will be entertained.
12. Purchaser reserves the right to change the quantity of stores allotted to them.
13. Purchaser reserves the right to get the manufacturing capacity of all the firms re-verified irrespective of the registration status or any other document/information.
14. Purchaser reserves the right to cancel /reject or scrap the tender, with no consequential liabilities.
15. **Two bid system;**  
All bidders are required to submit their offers in two covers online as under;-  
(A)**First cover (Technical Bid) is to be uploaded online and scanned copies in PDF format** should contain the following documents.
  - (i) The tenderer will clearly mention the name of the items for which he has quoted in the covering letter of the tender offer.
  - (ii) Tender documents (viz. Technical Compliance, brochures/catalogues of the quoted items containing Make/Model of the item. OEM/Authorization Certificate, registration/existence certificates of the firms/SSI Units and performance etc) duly page numbered completed and signed **BUT WITHOUT INDICATING THE RATES QUOTED.**
  - (iii) Scan copy of the Earnest money Deposit/CDR
  - (iv) Scan copy of the tender fee deposit/DD.
  - (v) Technical specifications of the items quoted with catalogues/brochures etc. and deposition of samples, (wherever asked).
  - (vi) Test report of each sample (wherever asked) shall be complete in all respects.
  - (vii) Tender with in-complete or without test report shall be rejected.
  - (viii) Sample test shall bear the seal of the testing laboratory prominently.
  - (ix) Registration Certificate/Industrial License of original manufacturer, with due validity for the items quoted.
  - (x) Certificate of being authorized dealer having proper and specific authorization (**as per format provided in appendix-1**) of the original manufacturer to quote against this tender.
  - (xi) Authorized dealer having authorization to quote against this tender, shall furnish his registration certificate with Central/ State Sale Tax Department for that particular item.
  - (xii) Scan copy of the NIT, all pages duly signed and stamped by the tenderer, in token of having understood and accepted the specifications, terms and conditions of the NIT and quoted items duly ticked.
  - (xiii) Performance statement/turn over for last 02 years as **appendix-III** of this NIT.
  - (xiv) Acceptance of the warranty/guarantee of the item.
  - (xv) Acceptance regarding delivery within the stipulated delivery schedule.
  - (xvi) PAN No./VAT/TIN No. self attested photocopies thereof be appended.
  - (xvii) Check list for tenders as per appendix-III.
  - (xviii) Any other relevant document which the firms wishes to submit


**B. SECOND COVER (COMMERCIAL/PRICE BID) SHOULD CONTAIN THE FOLLOWING:**

I.	Rate quoted by the firm should be on firm price basis as per bill of quantities (BOQs) of the e-tender to be downloaded from the website and Rates should be carefully quoted offline and then uploaded (Note): J&K Police does not issue road permits, concessional forms, state entry tax or octroi etc. exemption certificate, as such; Rate quoted shall be F.O.R consigne. <b>The zero basic cost quoted for any item in BOQ will be treated as nil quoted.</b>
II.	Rates quoted shall be final till the validity of the contract i.e, upto 30.12.2017 with a provision of further extension if finally approved by competent authority.

16. On due date of opening of tenders i.e . . . . .2016 at 1100 hours, technical bids only shall be opened and the technical evaluation process will start. Those who qualify the technical bid will be informed online with the announcement of financial bid opening date.
17. Before opening of financial bids, samples, if any, of the technically qualified bidders may be further physically evaluated by J&K Police. The samples not conforming to the functional requirements of the purchaser/tender requirements shall be rejected. Decision of Police department regarding acceptance/rejection of the samples shall be final.
18. **Security deposit.** The approved tenderers will have to supply the articles as per approved rate, approved sample and specification till the validity of the contract. They will have to furnish the **security deposit @ 5%** of the total value of the items contracted for in the shape of CDR/FDR issued by a nationalized/scheduled bank, at the time of signing of **Deed of Agreement** to ensure satisfactory performance of the equipment supplied, for warranty period from the date of acceptance of product. This security deposit will be released in due course after the expiry of specified warranty period/agreed contract. The charges on stamps etc. (cost of documents in totality) shall be borne by the supplier while executing the agreement. (No interest shall be paid by the purchaser on the security deposit). J&K Govt rules for security deposit in respect of local SSI Units shall be applicable.
19. **Guarantee/warranty.**  
Except as otherwise provided in the invitation to the tender, the contractor hereby declares that the goods, stores, articles sold/supplied to the purchaser under this contract shall be of the best quality and work man ship and new in all respects and shall be strictly in accordance with the specifications and particulars contained/mentioned in the contract. The contractor hereby guarantees that supplied goods/stores/articles would continue to conform to the specifications and quality for a period of 24 months from the date of acceptance. If during the period of 24 months supplied stores/good articles discovered not to be conforming to the specifications and quality or not giving satisfactory performance or have deteriorated, the decision of the purchaser in that behalf shall be final and binding on the contractor and purchaser shall be entitled to call upon the contractor to rectify/replace the goods/stores/articles or such portion thereof by the contractor and in such an event, the above mentioned warranty period shall apply to the goods/articles/stores rectified/replaced from the date of rectifications/replacement thereof. In case of failure of the contractor to rectify or replace with the goods etc. within specified time, the purchase shall be entitled to recover the cost will all expenses from the contractor for such defective stores by the available means.
20. **Liquidated Damages:** The successful tenderers shall have to make the supplies within the stipulated period as would be mentioned in the supply order. J&K Police reserve the right to impose liquidity damages of 5% of cost of un-supplied stores which shall depend upon the nature of supplies and extent of delay. However, period of delivery may be extended, at absolute discretion of J&K Police provided the delay is attributable to reasons beyond the control of the supplier but not attributable to business transactions and provided delay is communicated in writing justifying the cause of delay.
21. The supplies delivered by the successful tenderer shall be surveyed and verified by the J&K Police at Police Central Stores Jammu/Srinagar, as the case may be or after installation/commissioning of the same at the users end as the case may be.
22. PHQ shall have the discretion to send any number of samples from the received supplies from successful bidder for testing to one or more NABL/BIS accredited labs to check its conformity with the approved sample/specifications quoted in the NIT/supply order. In case of variation, the supply will be rejected at the cost and consequences of supplier.

23. As soon as the acceptance of the tender is communicated to the successful tenderers online, the contract shall be binding on him. The earnest money of the tenderer who backs out or withdraw his tender or fails to abide by it after acceptance thereof is communicated/posted to him shall be forfeited besides other penal remedies that may be available to the Govt. of the J&K under the law for the time being in force in the State of J&K;
24. The successful tenderer shall have to enter into an agreement on prescribed proforma (immediately after the issue of supply order) with the Department for due performance of the contract which shall be invariably furnished before the supply is made.
25. The tenderer is required to furnish an authenticated copy of Sales Tax/ VAT clearance certificates (alongwith the tender offer or at the time of payments) for the year 2015-16. The PHQ shall be at liberty to put to scrutiny, verification etc. the Sales tax /VAT clearance so furnished by the successful tenderer.
26. The authorities reserve absolute powers to reject or accept any tender or part thereof without assigning any reason and without any legal obligations. Any clause incorporated in the tender (not conforming to the NIT in any manner) shall make such an offer liable for rejection. Therefore, all the terms and conditions of the tender Notice including technical specifications shall be carefully studied for the sake of complete and comprehensive tender. Failure to comply with any of the conditions stipulated in the tender or instructions or the offer with insufficient particulars/documents shall lead to out-right rejection of the tender offer.
27. After the registration/placement of orders/execution of the agreement and formally depositing the security money, the tenderer backs out or fails to supply the goods, **the Security Deposit will be forfeited besides any other penal provisions, the department may decide as per norms;**
28. The formal deed incorporating the terms of the contract will be executed by the successful tenderer immediately after receipt of supply order. Failure to execute such bond/deed shall not however prevent the contract from being enforced against the tenderer. Any loss sustained by the Govt. as a result of re-tendering the contract shall be recovered from the defaulter, besides other penal provisions as the department may decide;
29. **Delivery of Goods**  
Goods shall be delivered (as per instructions in supply order) at Police Central Store, Gulshan Ground, Gandhi Nagar, Jammu and Police Central Store, Zewan, Srinagar as the case may be duly packed in perfect condition. The supplier if he so desires may insure the goods against loss or theft, destruction, damages by the natural calamity loss by exposure to weather or other wise viz war rebellion, riot etc. The insurance charges will have to be borne by the supplier and the department shall not pay such charges;
30. If the successful tenderer fails to supply the goods of the prescribed specification or fails to deliver the goods within the stipulated period as specifically mentioned in the supply order, the department shall be at liberty to arrange supplies by re-tendering or other-wise at the risk and cost of the tenderer. The department may give 15 days notice to the approved supplier in writing to make supplies of goods. In case of failure to supply the consignment on specified qualitative & quantitative terms, the department shall be lawfully entitled to forfeit the amount of the security money and take penal action to make good the loss sustained or excess cost incurred by the state in arranging of the supplies and other remedies that may be available to the department, under the law for the time being in force in the state. The imposable penalty can be 5% of the cost of goods depending upon nature of goods and cause of delay.
31. Either original manufacturer or his authorised dealer can participate for a particular item or sample(s) only. Tender offers where both original manufacturer and his authorised dealer participate for the same items shall be rejected.
32. If at any stage of tender process two or more bidders are found in collusion between/ among themselves with the intent to reduce competition, set price, share technical expertise, share business information/ decision, control/share production and or distribution of products and raise prices, all such tenders shall be out rightly rejected.
33. If at any stage of tender process any bidder/s is/are found influencing any Government Servant associated with or handling the tender process and supplies thereof, bid of such tenderer/s shall be rejected besides invoking any other provision available under rules.

34. All legal proceedings arising out of any dispute between the parties shall have to be settled in the Courts situated in Jammu and Srinagar and not elsewhere;
35. It shall be incumbent upon the supplier to provide entire supply, in conformity with the approved sample, strictly as per the specifications provided in the NIT/supply order. Supplies found deficient in quality/quantity viz-a-viz supply order shall be rejected. Penalty/damages would be imposed and recovered from the bidder as per terms and conditions indicated in the supply order.
36. The supplier shall not sublet the contract or any part thereof to any other agency.
37. The decision of the accepting authority shall be final as to the quality of stores and shall be binding on the supplier. In case the articles supplied are not according to the approved sample & specifications they shall be rejected and at our absolute discretion, rejected stores may be allowed to be made good by way of replacement. The loss caused due to rejection of the supplies shall be entirely borne by the supplier. The rejected articles shall be lifted by the supplier within one week from the date of rejection. The department shall in no case be responsible for any loss or damage that may occur to the rejected stores while these are in the premises of the purchaser.
38. The Department reserves the right to add or suitably modify the terms and conditions (without prior notification) as per exigencies and hence above shall not be construed as all inclusive.
39. Bidders submitting their bids on-line shall only be accepted and manual bids shall be rejected and no document other than mentioned in the NIT shall be accepted manually.
40. PHQ may call any bidder at any stage of the tender process for seeking any clarification.
41. Queries/ representations on tender enquiry received during pre-bid conference or within 10 days from the date of publication of this tender enquiry (whichever is earlier) would be entertained. All queries/representations received thereafter would not be entertained.
42. The bidder should furnish an undertaking APPENDIX-II of this NIT stating that he/ she is not connected in any manner with any other bidder in this tender with respect to manufacturing, distribution, supply, in sharing business information/documents price of the item etc. for the items quoted in this bid. If this is found to be incorrect, his/her tender would be rejected.
43. The detailed NIT, SBD and BOQ is available on J&K State e-Procurement portal [www.jktenders.gov.in](http://www.jktenders.gov.in). In case of any clarification, please contact on Phone no. **0194-2443031, 2443032**.

  
 (A/G Provision/Transport)  
 For Director General of Police,  
 J&K-Srinagar.  
 9/2/19



APPENDIX-I  
MANUFACTURER'S AUTHORIZATION LETTER/CERTIFICATE;

To,

The Director General of Police  
J&K, Srinagar

Subject:- \_\_\_\_\_

Sir,

Please refer to your e-tender enquiry (NIT) No- \_\_\_\_\_ of 2016 dated . . . . .2016.

1. We \_\_\_\_\_ are proven and reputable Manufacturers of \_\_\_\_\_ (name and description of the goods offered in the tender) having factories at \_\_\_\_\_ hereby authorize M/s \_\_\_\_\_ (name and address of the agent) to submit a tender. process the same further and enter into a contract with you against your requirement as contained in the above referred tender enquiry documents for the above said goods manufactured by us.
2. We also hereby extend our full warranty, CMC, support as per the conditions of contract, read with modifications, if any, in the special conditions of contract for the goods and services offered for supply by the above firm against this tender enquiry document.

Yours faithfully,

Signature with date, name and designation  
For and on behalf of M/s \_\_\_\_\_  
(Name and address of the Manufacturers)

Note:-

- a. This letter should be on the letter head of the manufacturing firm and should be signed by the person who is competent having the power of attorney.
- b. Scan copy of the original letter must be attached with tender documents.

SIGNATURE OF THE TENDERER.

APPENDIX-II

Reference e-NIT No. 28 of 2016 dated: 12 : 09: 2016

The Director General of Police,  
J&K, Srinagar.

Sir,

1. I \_\_\_\_\_ hereby state that I am not connected in any manner with any other bidder in this tender with respect to manufacturing distribution, supply, in sharing business information/documents and price of the items etc. quoted in this bid.
2. I further state that if this is found to be correct, the Police department is at liberty to reject my tender.
3. I further certify that my concern has neither been blacklisted nor banned by any central or state organizations to participate in tendering process.

Yours faithfully,

Signature with date, name and designation  
For and on behalf of M/s \_\_\_\_\_  
(Name and address of the manufacturer)

Note:-

- a. This undertaking should be on the letter head of the manufacturer firm and should be signed by the authorized person.
- b. Scan copy of the original letter must be attached with tender documents.

SIGNATURE OF THE TENDERER.

**APPENDIX-III**

Performance statement for last two years

Name of the firm	
Description of stores	
Quantity on order	
Value	
Original D.P	
Quantity supplied within original D.P	
Final Ext D.P.	
Last supply position.	
Reasons for delay in supplies ( if any)	

Signature of Tenderer

#### APPENDIX-IV

General instructions/CHECK LIST FOR TENDERERS;  
Before submission of tender documents, the tenderers should check whether they have complied with the following requirements or otherwise:-

S.No	Requirement to be checked before submission of the tender	Complied (Please indicate YES or NO after complying with the requirement.
1.	The tenderers for their own ease shall fill in relevant details of NIT offline & upload the same after ensuring accuracy	
2.	Whether tender fee scan copy is enclosed, with the technical bid?	
3.	Whether Earnest Money Deposit (EMD) scan copy is enclosed with the technical bid?	
4.	If an SSI Unit of J&K State, the same is indicated in the tender and scan copy of the valid registration/functional certificate is enclosed or not?	
5.	Whether list of the quoted items with name, make, model and catalogue/brochures of the quoted items are uploaded with technical bid?	
6.	Whether scanned copies of complete tender documents are up-loaded in the technical bid duly signed having official seal on all pages or not. However, the quoted items may also be ticked ?	
7.	Whether tender documents are submitted in two bid system i.e. Technical and Financial Bids separately as per tender enquiry or not?	
8.	Wherever asked to deposit the samples, whether sample conforming to the specifications has been deposited with Central Store Jammu/Srinagar (as asked for in the specifications of item)?	
9.	Whether offer validity as required in tender enquiry is accepted and indicated in your tender document or not?	
10.	Whether terms of delivery and period as required in tender enquiry is accepted and indicated in your tender document or not?	
11.	Whether payment terms as required in the tender enquiry is accepted and indicated in your tender document or not?	
12.	Whether compliance statement in the format as required in tender enquiry is enclosed with the supporting catalogue/ brochures?	
13.	Whether performance statement for last 2 years as required in tender enquiry in the format is enclosed or not. If not, reason thereof is given?	
14.	Whether warranty period as required in the tender enquiry is accepted or not?	
15.	Whether details of AMC condition after warranty period (wherever applicable) is included in the price bid or not, if required in the tender enquiry?	
16.	(a) Status of tenderer as to whether manufacturer or authorized agent of the manufacturer is indicated in the tender document. If authorised agent/dealer, valid authority letter for the stores offered from the manufacturer is enclosed or not. (b) If an Indian agent of foreign principal, whether the firm is enlisted with DGS&D under compulsory enlistment scheme or not. Copy of valid DGS&D enlistment certificate as Indian Agent of Foreign Principal is enclosed or not. If not, reason thereof is given in tender document or not?	

17.	If called for in the T.E. whether acceptance of functional demo/field trial clause, (whenever asked) is specified in your tender document or not?	
18.	If called for T.E. whether free training about use of equipment after supply, is accepted or no?	
19.	Whether indicated in the tender document, if your firm's business dealings is banned by any Govt/Private agency or recommended for blacklisting or not?	
20.	The complete commercial quote should be in Indian Rupees only. Currency of price should be clearly mentioned (wherever applicable).	
21.	Tenderers are required to indicate Basic Rate, Custom Duty, Excise Duty, Sales Tax/VAT and any other leviable taxes/charges separately in the relevant columns of the BOQ.	
22.	Entry Tax/Octroi/Toll Tax and other Statutory Taxes will be applicable at actual supported by documents, if admissible	
23.	Tenderer will attach justification of price quoted which should include the recently executed contracts of the store in question with Govt/Semi Govt Organizations etc.	
24.	No column should be left blank. Wherever amount is not quoted, the column should be clearly marked as 'Nil'.	

**List of sensitive and non sensitive documents.**

S. No	Sensitive Documents part of Technical bid	Non Sensitive Documents
I.	Valid Registration Certificate/Industrial License of original manufacturer for the items quoted	Performance statement of two years
II.	Valid authorization letter of original manufacturer in case of approved dealer	Govt. supply orders for same if any
III.	Valid registration certificate of approved dealer with Central/State Sales Tax Department for the items quoted	Income tax clearance
IV.	VAT clearance	
V.	SSI unit -registration certificate (local SSI units of J&K State only)	
VI.	SSI unit – functional certificate (local SSI units of J&K State only)	
VII.	Scanned copies of EMD and Tender Fee	
VIII.	Scanned copy of NIT duly Signed & sealed	

SIGNATURE OF THE TENDERER WITH SEAL

S.No.	Name of the item with specifications	Tentative quantity
1.	<p align="center"><b><u>SHORTWAVE THERAPY</u></b></p> <ul style="list-style-type: none"> <li>⇒ Continuous and pulse 1000 W with disk and pad electrodes.</li> <li>⇒ Input voltage: 220 – 240 V / 50 Hz.</li> <li>⇒ Output frequency: 27.12 MHZ</li> <li>⇒ Wave length: 11 M</li> <li>⇒ Power output: 500 watts cont. 1000 watt pulsed.</li> <li>⇒ Pulse rate: 20, 50, 100, 150, 200 Hz.</li> <li>⇒ Pulse width: 400 micro sec.</li> <li>⇒ Weight (kgs): 35 kgs Approximately.</li> </ul>	01 No.
2.	<p align="center"><b><u>MICROWAVE THERAPY UNIT</u></b></p> <ul style="list-style-type: none"> <li>⇒ 250 W (In constant mode) and 1600 W (in pulsating mode)</li> <li>⇒ Overheating Alarm</li> <li>⇒ Double protection system to guarantee extended magnetron life.</li> <li>⇒ Serial Controller: 16 bit microprocessor.</li> <li>⇒ Approx. 100 memorized protocols with space to store.</li> </ul>	01 No.
3.	<p align="center"><b><u>ULTRASOUND THERAPY UNIT</u></b></p> <p align="center"><b><u>(ADVANCED MODEL)</u></b></p> <ul style="list-style-type: none"> <li>⇒ Ultrasonic 1300 - 1/3 Mhz single channel unit with 5 cm<sup>2</sup> Ultrasound transducers.</li> <li>⇒ Ultrasonic 1500 – 1/3 Mh – 1/3 Mhz: Double Channel Unit with 1cm<sup>2</sup> / 5 cm<sup>2</sup> Ultrasound transducers.</li> <li>⇒ Knob regulated adjustable power from 0.1 to 3 W/cm<sup>2</sup> in continuous and pulse mode.</li> <li>⇒ 6 inch LCD Graphic display.</li> <li>⇒ Integrated serial controller.</li> <li>⇒ 16 bit microprocessor.</li> <li>⇒ Aprox. 50 memorized protocols with space for memorizing.</li> </ul> <p align="center"><b><u>(BASIC MODEL)</u></b></p> <ul style="list-style-type: none"> <li>⇒ Ultrasonic Frequency 1 MHz &amp; 3 MHz</li> <li>⇒ Ultrasonic output: Continuous 2.5 W/cm, Pulse 3.5/ cm.</li> <li>⇒ Pulse Frequency: 8 hz, 16 hz, 50 hz, 100 hz.</li> <li>⇒ Pulse Ratio: 1 : 2, 1 : 4, 1 : 7, 1 : 10, 1 : 16</li> <li>⇒ Treatment time: 0 – 99 min.</li> <li>⇒ Display: 16X2 LCD</li> <li>⇒ Program: 45 Pre-program (protocol) 90 Free prog.</li> </ul>	02 Nos.
4.	<p align="center"><b><u>MUSCLE STIMULATOR UNIT</u></b></p> <ul style="list-style-type: none"> <li>⇒ Plain Galvanic: For stimulation of Denervated muscle ion transfer. Maximum output 75 MA.</li> <li>⇒ Interrupted Galvanic Current: for diagnosis of muscle activity by plotting S-D Curve of defective muscle.</li> <li>⇒ Pulse width various in 10 steps from 0.01 to 300 MS.</li> <li>⇒ Faradic Current: Pulses generated are of short duration 0.6 MS @ 80 pulses/ second.</li> <li>⇒ Surged Faradic Current: Peak titanizing to be reached within 5 seconds. Rest period adjustable from 2 – 5 seconds.</li> </ul>	01 No.
5.	<p align="center"><b><u>CERVICAL &amp; LUMBER TRACTION UNIT</u></b></p> <ul style="list-style-type: none"> <li>⇒ Solid state circuitry.</li> <li>⇒ Patient Safety switch.</li> <li>⇒ Audio visual alarming system.</li> <li>⇒ Static intermittent switch.</li> <li>⇒ Digital Treatment timer.</li> <li>⇒ Hold time: 10, 20, 30, 40, 50, 60, 80 seconds.</li> <li>⇒ Rest: 1, 5, 10, 15, 20, 30 seconds.</li> <li>⇒ Weight: 15 Kgs.</li> <li>⇒ Consumption: 85 Watt</li> <li>⇒ Release Time: 1, 5, 10 Seconds.</li> <li>⇒ Traction Force: 5 to 45 Kgs with doubler 90 Kg.</li> <li>⇒ Voltage: 250V/ 50 Hz.</li> <li>⇒ Accessories: Cervical Collar, Spreader Bar, Pelvic Belt, Traction Doubler, Main Cord, Patient Safety Switch, Fuses.</li> </ul>	01 No.

	<p>⇒ 4 Fold Traction Bed: With features like Head section (lower) by 90° and lowers by 30° Leg section (lower) raises by 90° . Locking year facilities to be provided for table to be used as treatment couch so that lumbar section doesn't slide forward when the patient tries to sit/ sleep on it.</p>	
6.	<p style="text-align: center;"><b><u>4 CHANNEL TENS UNIT</u></b> <b><u>(DIGITAL)</u></b></p> <p>⇒ Output channel: 4 numbers.  ⇒ Frequency: 2, 40, 80, 130 Hz.  ⇒ Timer: 0 to 99 min.  ⇒ Display: Digital  ⇒ Mode Selection:  Burst: 7 Sec ON, 7 Sec OFF  Sweep: Impulses of all frequencies (2-40-80-130 Hz)  PWM: Variable pulse width: 150-250 Micro Seconds.  Wave Form: Monophasic.</p> <p style="text-align: center;"><b><u>(MANUAL)</u></b></p> <p>⇒ Output channel: 4 numbers.  ⇒ Modes: Continuous, Burst, Surge, Sweep, Mixed &amp; Random TENS.  ⇒ Display: Digital.  ⇒ Timer: 0-30 min.  ⇒ Wave form: Asymmetrical biphasic pulses.  ⇒ Intensity: 0-80mA.  ⇒ Pulse frequency: 2, 20, 40, 80, 120, 150 Hz (Selectable)  Input: 220-240 V/ 50 Hz.</p>	02 Nos.
7.	<p style="text-align: center;"><b><u>AUTOMATIC ORTHO OT TABLE</u></b></p> <p>⇒ Electro Hydraulic Operation with eight function remote Control table up and down, trendelenberg &amp; reverse trendelenberg, right lateral, left lateral, back rest, up and down.  ⇒ Remote control table with corded remote.  ⇒ Power supply should be 220 V, 5 amp and three pin plug point, Remote cord and main plug should come from head end of the Table i.e. from the bottom of the base to facilitate the anesthetist.  ⇒ Table should have 100mm central lock castors for longitudinal and lateral movements, the castors should be made of polyurethane and molded to prevent water and rust entering inside.  ⇒ Braking should be effective and have a dead lock when broken: it should be on the head and side of the table, to facilitate the anesthetist and one brake on the tail end for better stability.  ⇒ Leg should be two pieces right and left and should be detachable and about able and also should move up &amp; down  ⇒ Table top should have five sections and should be breakable into inverted V shape from the head side.  ⇒ Trendelenberg and lateral mechanism should be concealed with bellows column should be straight without any projection on right or left.  ⇒ Table column size should be 280mmx280mm.  ⇒ Table should have provision to view spine up and lateral C - Arm.  ⇒ Polyurethane detachable cushion top should be provided on the table top.  ⇒ Base cover should be covered with impact, shock resistant, fire resistant and disinfectant free material.  ⇒ Base should be border in the head and the column casing. Table top frame, traction bars and all accessories made of non-corrosive steel. The main column should be covered with non-corrosive stainless steel.  ⇒ Inbuilt battery backup should be provided.  ⇒ The table top should be reversible, the head rest attachment should be fixed to the tail side and the tail side attachments should be fixed to the head rest.  ⇒ The table should have provision to fix top leg traction attachment.  ⇒ The table should have manual persecutor for all positions.  ⇒ In case of electro components failure, the table should be operated manually by mechanical pedaling for all positions.  ⇒ The table should be capable of taking off centered load from neck to toe and screened into C arm without any disturbance.  ⇒ ISO 9001, ISO 13485 approved.  ⇒ The entire vertebral column should be viewed without any hindrance.  ⇒ Patient sitting position, with Trendelenberg maximum should be possible.</p>	01 No.

	<p>⇒ Minimum height 27" preferred.</p> <p>⇒ Should adapt any field sugitha and lyela retractor.</p> <p>Should have attachments:</p> <ul style="list-style-type: none"> <li>• Top leg Att for neck femur surgery.</li> <li>• "C" arm compatible hanging tibial support.</li> <li>• Horse shoe head rest/ Horse shoe head rest support.</li> <li>• Steimen pin holder.</li> <li>• Tibia condyl fixation traction support.</li> <li>• Femur interlocking lateral position accessories.</li> <li>• Hand surgery board /Finger traction board.</li> <li>• Knee chest support.</li> <li>• Gas spring leg bed</li> <li>• Perineal post.</li> </ul>	
8.	<p style="text-align: center;"><b><u>Micro Laryngeal Surgery (Direct Laryngoscope)</u></b></p> <ol style="list-style-type: none"> <li>1) Laryngoscope. Kleinsasser for Microsurgery-Adult-Large Proximal end approximately 28-30mmWx18.5mmH/Distal end 28.5mmW.x18.5mmH/Distal end i15mm. Length 18-20cm.</li> <li>2) Laryngoscope. Kleinsasser for Microsurgery-Adult-Medium-Proximal end approximately 25-27mmWx15.5mmH/Distal end i13.5mm. Length 18-20cmcm.</li> <li>3) Laryngoscope. Kleinsasser for Microsurgery-Child-Proximal end 20.5mmWx13mmH/Distal end i10.5mm. Length 15cm.</li> <li>4) Laryngoscope Tube, Anterior Commissure,Jackson, Adult. OD 14-16mm. Length 18-20cm.</li> <li>5) Laryngoscope Tube, Anterior Commissure,Jackson, Child. OD 12-13mm. Length 14-16cm.</li> <li>6) Fibre Light Carrier, for J/AC Laryngoscope. Adult.OD 2.5mm. Length approximately 16-18cm.</li> <li>7) Fibre Light Carrier, for J/AC Laryngoscope. Child . OD 2.5mm. Length approximately 12cm.</li> <li>8) Knife, Micro-Laryngeal. Sickle. Str.Length 23cm.</li> <li>9) Knife, Micro-Laryngeal. Plester.Length 23cm.</li> <li>10) Knife, Micro-Laryngeal. Oval.Str..Length 23cm</li> <li>11) Handle for use with straight instruments</li> <li>12) Suction Tube, Rigid, with stilette. Length 25cm. OD 2.5mm.</li> <li>13) Forceps(Fcps)., Micro-Laryngeal, 2mm. Cup jaw. Straight.Length 23cm.</li> <li>14) Fcps., Micro-Laryngeal, 2mm. Cup jaw. Rt Length 23cm</li> <li>15) Fcps., Micro-Laryngeal, 2mm. Cup jaw. Lt.Length 23cm.</li> <li>16) Fcps., Micro-Laryngeal. Serrated., Str. Length 23cm.</li> <li>17) Fcps., Micro-Laryngeal, 3x4mm Oval Cup.patterson/Kleinsasser. Length 23cm.</li> <li>18) Insulated Cannula for suction and coagulation O.D. 3mm working length26cm</li> </ol>	01 No.
9.	<p><b><u>SUCTION APPARATUS</u></b></p> <ol style="list-style-type: none"> <li>1. Should be approved from Bureau of Indian Standards.</li> <li>2. Anti-corrosive. Powder Coated Chassis</li> <li>3. 200 Watt Heavy Thermal Protector Electric Motor</li> <li>4. 2 X 2.0 Ltrs. Polycarbonate jars with overflow safety</li> <li>5. Noise level of suction apparatus from is 50 dB +/- 03 dB</li> <li>6. Electrical Requirement - 220-230V, 50Hz, 1Phase</li> </ol>	04 Nos.




	<p>7. Ideal for MTP / medical / surgical procedures</p> <p>8. Oil Free Rocker Piston Pump with -750 +/- 10 mmHg</p> <p>9. Free air displacement &gt;100 LPM</p> <p>10. Non collapsible Silicon tubing</p> <p>11. Bacterial filter fitted on top</p> <p>12. Heavy duty HN-65 PU Castors</p> <p>13. 100 mm Vacuum Gauge</p> <p>14. Minimum Noise production.</p>															
10.	<p><b>CAUTERY</b></p> <ul style="list-style-type: none"> <li>● Electrosurgical Generator with Instant Response Technology designed for all general surgical, urological, laparoscopic and orthopaedic procedures with following accessories: <ul style="list-style-type: none"> <li>○ Monopolar foot switch 3 mt cord.</li> <li>○ Foot switch</li> <li>○ Disposable Hands with pencil.</li> <li>○ Polhesive II ret elec.</li> <li>○ Universal active adaptor.</li> <li>○ BIP Cord.</li> <li>○ Cased Disposable Bipolar Cord.</li> <li>○ With CVT compatible with cautery.</li> <li>○ Cautery stand.</li> </ul> </li> </ul>	01 No.														
11.	<p style="text-align: center;"><b><u>PHYSIO PARAFFIN WAX BATH</u></b></p> <p>Physio Paraffin Wax Bath, tank is made of heavy stainless steel used and is mounted on ball bearing rubber castors, for easy mobility. Works in electricity 220 – 230 volts and fitted with two special tanks is impression type heaters. Thermostatically controlled comes complete with three core wire pin plug:</p> <p><b>Inner tank size:</b></p> <ul style="list-style-type: none"> <li>● Large Size: 18" X 16" X 10"</li> <li>● Medium Size: 14" X 09" X 7"</li> <li>● Small Size: 14" X 7" X 7"</li> </ul>	01 No.														
12.	<p style="text-align: center;"><b><u>OPERATING MICROSCOPE (OPHTHALMOLOGY)</u></b></p> <p>⇒ Optics:</p> <ul style="list-style-type: none"> <li>○ Apochromatic corrected optics with objective F = 200 mm.</li> <li>○ Motorized zoom type magnification changer.</li> <li>○ Titable binocular tubes (0-90°)</li> <li>○ Focus range: 30- 45 mm.</li> <li>○ Integrated assistant interface.</li> </ul> <p>⇒ Illumination:</p> <ul style="list-style-type: none"> <li>○ LED Light Source.</li> <li>○ Filter: Heat absorbing, Blue blocking, colour enhancement. Retinal protection.</li> </ul> <p>⇒ Operation:</p> <ul style="list-style-type: none"> <li>○ Floor Stand with lockable wheels – cataract version.</li> <li>○ Motorised fine focus, X-Y movement.</li> <li>Continuous Zoom Magnification from foot switch.</li> </ul> <p>⇒ Optional Accessories:</p> <ul style="list-style-type: none"> <li>○ Beam Splitter with Camera and C-mount.</li> <li>○ Video recording facility.</li> </ul>	01 No.														
13.	<p><b><u>Binocular Microscope.</u></b></p> <table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 50%;">1. Optic system:</td> <td>UIS 2 (Universal Infinity System)</td> </tr> <tr> <td>2. Illumination System:</td> <td>Built in Transmitted illumination system, LED</td> </tr> <tr> <td>3. Focusing:</td> <td>Stage height movement (coarse movement stroke 20 mm)</td> </tr> <tr> <td>4. Revolving nose piece:</td> <td>Fixed quadruple nosepiece</td> </tr> <tr> <td>5. Stage:</td> <td>Wire movement mechanical fixed stage 120x132 mmtravelling range: 76mm(X) x 30 mm (Y)</td> </tr> <tr> <td>6. Observation tube:</td> <td>30° inclined binocular tube</td> </tr> <tr> <td>7. Condenser:</td> <td>Abbe type with aperture iris diaphragm N.A. 125</td> </tr> </table>	1. Optic system:	UIS 2 (Universal Infinity System)	2. Illumination System:	Built in Transmitted illumination system, LED	3. Focusing:	Stage height movement (coarse movement stroke 20 mm)	4. Revolving nose piece:	Fixed quadruple nosepiece	5. Stage:	Wire movement mechanical fixed stage 120x132 mmtravelling range: 76mm(X) x 30 mm (Y)	6. Observation tube:	30° inclined binocular tube	7. Condenser:	Abbe type with aperture iris diaphragm N.A. 125	01 No.
1. Optic system:	UIS 2 (Universal Infinity System)															
2. Illumination System:	Built in Transmitted illumination system, LED															
3. Focusing:	Stage height movement (coarse movement stroke 20 mm)															
4. Revolving nose piece:	Fixed quadruple nosepiece															
5. Stage:	Wire movement mechanical fixed stage 120x132 mmtravelling range: 76mm(X) x 30 mm (Y)															
6. Observation tube:	30° inclined binocular tube															
7. Condenser:	Abbe type with aperture iris diaphragm N.A. 125															

	<p>8. Objectives: Plan achromatic objective (Anti-fungal)  4X N.A.:0.10 W.D: 18.5 mm  10X N.A: 0.25 W.D: 10.6mm  40X N.A: 0.65 W.D.:0.06 mm  100XOil: N.A.: 1.25 W.D: 0.13mm</p> <p>9. Eye piece (10X): Field number (F.N), 20 (Anti-Fungus)</p> <p>10. Optional accessories: Mirror unit, 15X eye piece (F.N. 12, anti fungus), micrometer, filter holder, dark field stop, micro image projection system, phase-contrast attachment/ 10X, 40X, simple Polarizing attachment.</p> <p>11. Weight: Approx. 6 Kgs</p> <p>12. Rated voltage/electric current: AC 100-240 V, 50/60 Hz 0.4 A</p> <p>13. Power consumption: 17w</p> <p>14. Comprehensive maintenance contract after expiry of warranty</p>																					
14.	<p><b>PHACO EMULSIFICATION MACHINE</b></p> <ul style="list-style-type: none"> <li>⇒ 4 to 6 Piezo-electric crystals U/S Titanium Handpiece with Frequency of approx. 40 KHz.</li> <li>⇒ U/S Delivery modes: Continuous, Pulse, Burst and Programmable emission mode (adjustable duty cycle).</li> <li>⇒ Continuous tip stroke length control and adjustment for consistent U/S power delivery.</li> <li>⇒ Range of Phaco tips like straight, Bent and flared tips.</li> <li>⇒ MICS Phaco tips for 1.8 mm incision co-axial phaco.</li> <li>⇒ Inbuilt Bipolar Diathermy with 2 MHz Frequency, 7 W power.</li> <li>⇒ Double pump (Venturi and peristaltic) aspiration system with vacuum range of 5 to 500 mmHg and Flow rate 2 to 50 cc/ min with adjustable aspiration rise time.</li> <li>⇒ Autoclavable I/A tubing with facility of Anterior chamber surge protection system.</li> <li>⇒ Vitrectomy cutting rate ranging from 60 to 4000 cuts per minute with Pneumatically powered Guillotine Cutter.</li> <li>⇒ Facility of upgrading to posterior vitrectomy unit.</li> <li>⇒ Min 12 Inches LCD with touch screen and user friendly control software.</li> <li>⇒ Programmable dual linear foot switch for linear/ panel control of all equipment functions.</li> <li>⇒ More than 1000 user programs saving facility.</li> <li>⇒ European CE Approved.</li> <li>⇒ Phaco Trolley with Manually adjustable I/V Stand.</li> <li>⇒ Suitable UPS.</li> <li>⇒ Accessories: <ul style="list-style-type: none"> <li>○ U/S Hand pieces.</li> <li>○ Co-axial I/A Hand pieces.</li> <li>○ Bimanual I/A Hand pieces.</li> <li>○ Phaco Tips.</li> <li>○ Silicon Sleeves.</li> <li>○ Test Chambers.</li> <li>○ I/A Tubings.</li> <li>○ Diathermy Forceps.</li> <li>○ Diathermy Cable.</li> <li>○ Tip Wrench.</li> <li>○ Sterilization Tray.</li> <li>○ Vitrectomy cutters.</li> </ul> </li> </ul>	01 No.																				
15.	<p><b>GENERAL SURGERY AND ALLIED INSTRUMENTS</b></p> <table border="1" data-bbox="188 1641 1353 1991"> <tr> <td>1. Sponge holder Forceps (assorted size):</td> <td>02 dozen</td> </tr> <tr> <td>2. Artery forceps St. 6":</td> <td>04 dozen</td> </tr> <tr> <td>3. Artery forceps Curved 6":</td> <td>04 dozen</td> </tr> <tr> <td>4. Mosquito Artery Forcep St.5"</td> <td>02 dozen</td> </tr> <tr> <td>5. Mosquito Artery Forcep St. 6"</td> <td>02 dozen</td> </tr> <tr> <td>6. Mosquito Artery Forcep St. 8"</td> <td>02 dozen</td> </tr> <tr> <td>7. Mosquito Artery Forcep Curved 5"</td> <td>02 dozen</td> </tr> <tr> <td>8. Mosquito Artery Forcep Curved 5"</td> <td>02 dozen</td> </tr> <tr> <td>9. Mosquito Artery Forcep Curved 8"</td> <td>02 dozen</td> </tr> <tr> <td>10. Scissors 6" St. (Mayo)</td> <td>02 dozen.</td> </tr> </table>	1. Sponge holder Forceps (assorted size):	02 dozen	2. Artery forceps St. 6":	04 dozen	3. Artery forceps Curved 6":	04 dozen	4. Mosquito Artery Forcep St.5"	02 dozen	5. Mosquito Artery Forcep St. 6"	02 dozen	6. Mosquito Artery Forcep St. 8"	02 dozen	7. Mosquito Artery Forcep Curved 5"	02 dozen	8. Mosquito Artery Forcep Curved 5"	02 dozen	9. Mosquito Artery Forcep Curved 8"	02 dozen	10. Scissors 6" St. (Mayo)	02 dozen.	
1. Sponge holder Forceps (assorted size):	02 dozen																					
2. Artery forceps St. 6":	04 dozen																					
3. Artery forceps Curved 6":	04 dozen																					
4. Mosquito Artery Forcep St.5"	02 dozen																					
5. Mosquito Artery Forcep St. 6"	02 dozen																					
6. Mosquito Artery Forcep St. 8"	02 dozen																					
7. Mosquito Artery Forcep Curved 5"	02 dozen																					
8. Mosquito Artery Forcep Curved 5"	02 dozen																					
9. Mosquito Artery Forcep Curved 8"	02 dozen																					
10. Scissors 6" St. (Mayo)	02 dozen.																					

11. Scissors 6" Curved (Mayo)	02 dozen
12. Scissors 8" St.(Metzenbam):	02 dozen.
13. Scissors 8" Curved (Metzenbam):	02 dozen.
14. Pick up forceps tooth 8": (Thumb forecep)	02 dozen
15. Pick up forceps Non tooth 8" (Thumb forecep)	02 dozen.
16. Adison forceps Tooth (assorted size):	02 dozen
17. Adison forceps non tooth (assorted size):	02 dozen
18. Micro scissors St	01 dozen
19. Micro scissors Curved:	01 dozen
20. Cat's Paw Retractor:	06 No.s
21. Ellies Forceps (6"):	02 dozen
22. Babcock Forceps (Assorted size):	01 dozen
23. SPC Trocar:	02 No.s
24. NSV Set:	02 No.s
25. Myomectomy screw:	01 No.
26. Sterilizing forcep 8" Cheattle:	02 No.
27. Needle Holder4" Mayo Hegar 4"	02 No.s.
28. Needle Holder4" Mayo Hegar (5"	02 No.s
29. Needle Holder4" Mayo Hegar 7"	02 No.s
<b><u>MASTOID AND TYMPANOPLASTY INSTRUMENTS</u></b>	
1. Periosteum elevator, Freer's -	1 Set
2. Hartmann aural forceps -	2No's
3. CHISEL (JANSEN) 140mm,( 2,mm)	01 No.
4. CHISEL (JANSEN) 140mm .3,mm)	01 No.
5. CHISEL (JANSEN) 140mm .4,mm)	01 No.
6. OSTEOTOME (JANSEN) 140mm,( 2,mm)	01 No.
7. OSTEOTOME (JANSEN) 140mm 3,mm)	01 No.
8. OSTEOTOME (JANSEN) 140mm 6,mm)	01 No.
9. OSTEOTOME (JANSEN) 140mm 8,mm)	01 No.
10. GAUGE (JENKINS) 140mm 3, mm)	01 No.
11. GAUGE (JENKINS) 140mm 4,mm)	01 No.
12. GAUGE (JENKINS) 140mm 6,mm)	01 No.
13. GAUGE (JENKINS) 140mm 8,mm)	01 No.
14. Lempert nibbling forceps, curved -	1No,
15. Lempart endaural speculum ( medium,	1No,
16. Lempart endaural speculum ( large)	1No,
17. Fine granulation tissue forceps -	1No,
18. Mastoid curette (Lempert) - 2mm,	1No
19. Mastoid curette (Lempert) -, 4mm,	1No
20. Mastoid curette (Lempert) -, 5mm	1No
21. Mastoid curette (Lempert) - 6mm,	1No,
22. Lake Seeker -	1No,
23. Micro Aural Crocodile Forceps (TC)Serrated, Fine Straight -	2N0'S,
24. Micro Aural Crocodile Forceps (TC)Serrated, Fine Upward -	2 No'S,
25. Micro Aural Cup Forceps(TC), Fine Straight -	1No,
26. Micro Aural Cup Forceps(TC), Fine Upward -	1No,
27. Micro Aural Scissors(TC), Fine Straight -	1No.
28. Micro Aural Scissors(TC), Fine Upward -	1No.
29. Curette (House) double ended -	1,No
30. Pick 45°, length 16 cm, size 1 mm -	1No
31. Same, size 1.5 mm -	1No
32. Pick 90°, length 16 cm, size 1 mm, -	1No
33. Same, size 1.5 mm -	1No
34. Same, size 2 mm -	1 No
35. Manual Perforator, diameter 0.4 mm, length 16 cm -	1No
36. Same, diameter 0.5 mm-	1No

37. Same, diameter 0.6 mm! -	1No
<b>INSTRUMENTS FOR MICROLARYNGEAL SURGERY</b>	
1. Teeth protector, it should be made up of silicon and should be autoclavable, -	1No
2. Needle holder, it should be with ratchet, working length approx. 23cm, -	1No
3. Laryngeal biopsy forceps, it should be heavy, approx. size 3x4mm, working length should be 23cm, -	1No
4. Scissors, straight, working length should be approx.23cm, -	1No
5. Scissors, it should be angled 45 deg curved to left, working length approx. 23cm -	1 No
6. Alligator forceps, serrated, straight working length approx. 23cm, -	2No's
7. Cutting forceps, straight, curved to upward, curved to left, curved to right, with approx. 2mm cupped jaws, working length approx. 23cm	01 No
8. Hook, blunt and sharp, with probe button, working length approx. 23cm	1No,
9. Suction tube, outer diameter should be around 2.5mm, -	1No
<b>NASAL AND FESS INSTRUMENTS</b>	
1. Skin hook (Gillies) TITANIUM -	2No'S,
2. ALAR Retractor -	2No'S,
3. FREER's Elevator -	1No,
4. TILLEY LICHTWITZ Antral Trocar & cannula -	1No,
5. FOMAN Angular Scissors (TC) -	1No,
6. Nasal Polypus Forceps (WATSON WILLIAMS) -	1No,
7. Nasal Biopsy Punch Forceps -	1No,
8. Nasal Gouges - 2mm, 4mm, 6mm, 8mm, -	1No,
9. Turbinectomy scissors-	1No
10. Cottle Retractor blunt -	2No'S,
11. Antrum curette (19cm) - Forward cutting - 1NO, Backward cutting -	1No,
12. Through cutting punch, straight - 1NO, up 45° -	1No,
13. LUC's Forceps, Large -	1No,
14. Sickle knife, pointed, 19cm long, -	1No,
15. Conical suction tube should be malleable, with finger grip plate, luer-lock, Outer diameter 2.5cm, working length 13cm -	1No,
16. V. Eicken antrum cannula, luer-lock, long curved, malleable, outer diameter 3mm & 2.5mm, length 12.5cm	1No,
17. OD V Eicken antrum cannula, Luer-lock, with cut-off hole, short curved, outer diameter 3mm & 4 mm, length 12.5cm	1No,
18. BLAKESLEY forceps St. 13 cm size 1,	1No
19. BLAKESLEY forceps St. 13 cm size 1	1No
20. BLAKSELEY forceps Angled 45° size 1	1No
21. BLAKSELEY forceps Angled 45° size 2	1No
22. BLAKSELEY forceps angled 90° upturned size 2 -	1No
23. Ballenger chisel 4mm wide v shaped 5-6'' -	1No,
24. Ballenger chisel 6mm wide v shaped 6-7'' -	1No,
25. Kerrison Costen Rongeur Bite 4x9 mm 90° curved 85 mm shaft -	1No,
26. Struycken Rhinoforce nasal cutting forceps, working length 13cm -	1No,
27. Rhinoforce nasal scissors, Straight, working length 13cm -	1No,
<b>INSTRUMENTS FOR TONSILLECTOMY AND ADENOIDECTOMY</b>	
1. Doughty slotted mouth gag with 5 tongue plates	1No,
2. Negus tonsil artery forceps - Half curved -	1No,
3. Negus tonsil artery forceps - Full curved -	1No,
4. BIRKETT Tonsil artery forceps - Curved -	1No,
5. Tonsil Pillar Retractor & Dissector -	2No's,
6. Tonsil Knife -	1NO

  
 For Director General of Police,  
 J&K, Srinagar